



# CERTIFICATE OF ACCREDITATION

**The ANSI National Accreditation Board**

Hereby attests that

**Techmaster Electronics, LLC**  
**1070 Joshua Way**  
**Vista, CA 92081**  
(and satellite locations as listed on the scope)

Fulfills the requirements of

**ISO/IEC 17025:2017**

and national standard

**ANSI/NCSL Z540-1-1994 (R2002)**

In the field of

**CALIBRATION, DIMENSIONAL MEASUREMENT AND TESTING**

This certificate is valid only when accompanied by a current scope of accreditation document.  
The current scope of accreditation can be verified at [www.anab.org](http://www.anab.org).

Jason Stine, Vice President

Expiry Date: 29 October 2024

Certificate Number: AC-1736



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.  
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory  
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



**SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017  
AND  
ANSI/NCSL Z540-1-1994 (R2002)**

**Techmaster Electronics, LLC**

1070 Joshua Way  
Vista, CA 92081

Ernesto Matamoros Phone: 760-536-0227 ext. 405  
Email: [quality.mx@techmaster.us](mailto:quality.mx@techmaster.us) Website: [www.techmaster.us](http://www.techmaster.us)

Services performed at satellite locations as indicated  
46782 Lakeview Blvd. Fremont, CA 94538  
6120 Hanging Moss Rd. Orlando, FL 32807

**CALIBRATION, DIMENSIONAL MEASUREMENT AND TESTING**

Valid to: **October 29, 2024**

Certificate Number: **AC-1736**

**CALIBRATION**

**Acoustics and Vibration**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Vibration	(0.2 to 36) gpk (1 to 20) Hz (20 to 100) Hz (100 to 2 500) Hz (2500 to 10 000) Hz	3.1 % of reading + 0.04 g 3.1 % of reading + 0.04 g 3.1 % of reading + 0.04 g 3.4 % of reading + 0.04 g	Vibration Calibrator System  Fremont, CA
Sound Level Meters	(94, 114) dB @1 kHz	0.28 dB	Sound Calibrator  Vista, CA

### Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH - Source	4.01 pH 7.00 PH 10.00 pH	0.017 pH 0.018 pH 0.031 pH	Standard Solutions  Vista, CA Fremont, CA Orlando, FL
Conductivity - Source	(100 to 10 000) $\mu\text{S/cm}$	1 % of calibrated value	Standard Solutions  Vista, CA Fremont, CA Orlando, FL

### Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Voltage – Source <sup>1</sup>	Up to 220 mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V 220 V to 1.1 kV	6.8 $\mu\text{V/V} + 0.48 \mu\text{V}$ 4.6 $\mu\text{V/V} + 0.8 \mu\text{V}$ 3.2 $\mu\text{V/V} + 3.2 \mu\text{V}$ 2.4 $\mu\text{V/V} + 27 \mu\text{V}$ 4.7 $\mu\text{V/V} + 43 \mu\text{V}$ 6 $\mu\text{V/V} + 0.85 \text{ mV}$	High Performance Multifunction Calibrator  Vista, CA Orlando, FL
DC Voltage – Source	Up to 220 mV 220 mV to 2.2 V (2.2 to 11) V (11 to 22) V (22 to 220) V 220 V to 1.1 kV	7.5 $\mu\text{V/V} + 0.4 \mu\text{V}$ 5 $\mu\text{V/V} + 1 \mu\text{V}$ 3.5 $\mu\text{V/V} + 3.1 \mu\text{V}$ 3.5 $\mu\text{V/V} + 8.1 \mu\text{V}$ 5 $\mu\text{V/V} + 83 \mu\text{V}$ 6.5 $\mu\text{V/V} + 0.68 \text{ mV}$	High Performance Multifunction Calibrator  Fremont, CA
DC Voltage – Source <sup>1</sup> Fixed Points	1 V 1.018 V 10 V	14 $\mu\text{V}$ 17 $\mu\text{V}$ 35 $\mu\text{V}$	DC Reference Standard  Vista, CA
DC Voltage – Measure <sup>1</sup>	Up to 100 mV 100 mV to 1 V (1 to 10) V (10 to 100) V 100 V to 1 kV	7.2 $\mu\text{V/V} + 0.55 \mu\text{V}$ 7 $\mu\text{V/V} + 0.42 \mu\text{V}$ 6.9 $\mu\text{V/V} + 0.86 \mu\text{V}$ 9.2 $\mu\text{V/V} + 38 \mu\text{V}$ 9.3 $\mu\text{V/V} + 0.13 \text{ mV}$	Precision 8.5 Digit Multimeter  Fremont, CA Orlando, FL
	(2 to 200) mV 200 mV to 2 V (2 to 20) V (20 to 200) V 200 V to 1 kV	6.7 $\mu\text{V/V} + 0.2 \mu\text{V}$ 4.3 $\mu\text{V/V} + 0.5 \mu\text{V}$ 4.3 $\mu\text{V/V} + 4.8 \mu\text{V}$ 6.7 $\mu\text{V/V} + 98 \mu\text{V}$ 6.7 $\mu\text{V} + 0.63 \text{ mV}$	Precision 8.5 Digit Multimeter  Vista CA

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC High Voltage - Measure <sup>1</sup>	Up to 10 kV (10 to 100) kV	0.35 mV/V + 0.09 V 0.63 mV/V + 4.1 V	High Voltage Meter  Vista, CA Fremont, CA Orlando, FL
DC Current – Source <sup>1</sup>	Up to 220 $\mu$ A 220 $\mu$ A to 2.2 mA (2.2 to 22) mA (22 to 220) mA 220 mA to 2.2 A (2.2 to 11) A	0.003 7 % of reading + 5.8 nA 0.003 1 % of reading + 7.4 nA 0.003 1 % of reading + 44 nA 0.004 % of reading + 0.69 $\mu$ A 72 $\mu$ A/A + 14 $\mu$ A 0.023 % of reading + 0.47 mA	High Performance Multifunction Calibrator w/ Transconductance Amplifier  Vista, CA Orlando, FL
DC Current – Source <sup>1</sup>	Up to 220 $\mu$ A 220 $\mu$ A to 2.2 mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	40 $\mu$ A/A + 6 nA 35 $\mu$ A/A + 7.6 nA 35 $\mu$ A/A + 48 nA 55 $\mu$ A/A + 0.74 $\mu$ A 0.13 mA/A + 13 $\mu$ A	High Performance Multifunction Calibrator  Fremont, CA
DC Current – Source <sup>1</sup>	330 mA to 1.1 A (1.1 to 3) A (3 to 11) A (11 to 20) A	0.15 mA/A + 44 $\mu$ A 0.27 mA/A + 0.13 mA 0.39 mA/A + 0.41 mA 0.79 mA/A + 0.41 mA	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
DC Current – Source <sup>1</sup> Clamp-On Meters	(10 to 16.5) A (16.5 to 150) A (150 to 1 025) A	0.5 % of reading + 0.55 A 0.5 % of reading + 0.57 A 0.5 % of reading + 0.74 A	Multi Product Calibrator w/ Current Coil  Vista, CA Fremont, CA Orlando, FL
DC Power – Source <sup>1</sup> 33 mV to 1 020 V 329.99 mA range 2.999 9 A range 20.5 A range	10 $\mu$ W to 330 W 10 mW to 3 kW 100 mW to 21 kW	0.01 % of reading + 1.2 nW 0.04 % of reading 0.1 % of reading	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
DC Power - Measure	Up to 10 W (>10 to 15) W (>15 to 100) W (>100 to 150) W	0.043 % of reading + 0.074 W 0.043 % of reading + 0.015 W 0.043 % of reading + 0.098 W 0.043 % of reading + 0.15 W	Precision Power Analyzer  Vista, CA

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Measure <sup>1</sup>	Up to 100 nA 100 nA to 1 µA (1 to 10) µA (10 to 100) µA 100 µA to 1 mA (1 to 10) mA (10 to 100) mA 100 mA to 1 A	22 µA/A + 93 pA 18 µA/A + 0.1 nA 28 µA/A + 0.14 nA 30 µA/A + 0.98 nA 30 µA/A + 6.5 nA 30 µA/A + 63 nA 44 µA/A + 1 µA 0.14 mA/A + 12 µA	Precision 8.5 Digit Multimeter  Fremont, CA Orlando, FL
DC Current – Measure <sup>1</sup>	(2 to 200) µA (200 µA to 2) mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A	15 µA/A + 0.49 nA 15 µA/A + 4.8 nA 17 µA/A + 48 nA 59 µA/A + 0.96 µA 0.23 mA/A + 25 µA 0.49 mA/A + 0.36 mA	Precision 8.5 DigitMultimeter  Vista, CA
DC Current – Measure <sup>1</sup>	(1 to 100) A (100 to 600) A (600 to 2 000) A	85 µA/A 127 µA/A 0.29 % of reading + 0.02 A	Precision 8.5 Digit Multimeter w/ Current Shunts  Vista, CA Fremont, CA Orlando, FL
AC Voltage – Source <sup>1</sup>	(1 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.013 % of reading + 7.2 µV 0.008 3 % of reading + 4.2 µV 0.007 2 % of reading + 4.2 µV 0.016 % of reading + 4.6 µV 0.042 % of reading + 5.3 µV 0.095 % of reading + 10 µV 0.13 % of reading + 20 µV 0.25 % of reading + 21 µV  0.021 % of reading + 4.9 µV 0.008 3 % of reading + 4.3 µV 0.007 % of reading + 4.4 µV 0.017 % of reading + 4.8 µV 0.042 % of reading + 6.1 µV 0.092 % of reading + 12 µV 0.13 % of reading + 22 µV 0.22 % of reading + 33 µV	High Performance Multifunction Calibrator w/ Transconductance Amplifier  Vista, CA Orlando, FL



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(22 to 220) mV		High Performance Multifunction Calibrator w/ Transconductance Amplifier  Vista, CA Orlando, FL
	(10 to 20) Hz	0.022 % of reading + 16 $\mu$ V	
	(20 to 40) Hz	0.006 4 % of reading + 15 $\mu$ V	
	40 Hz to 20 kHz	0.004 4 % of reading + 10 $\mu$ V	
	(20 to 50) kHz	0.009 % of reading + 14 $\mu$ V	
	(50 to 100) kHz	0.028 % of reading + 23 $\mu$ V	
	(100 to 300) kHz	0.056 % of reading + 35 $\mu$ V	
	(300 to 500) kHz	0.13 % of reading + 39 $\mu$ V	
	500 kHz to 1 MHz	0.23 % of reading + 0.12 mV	
	220 mV to 2.2 V		
	(10 to 20) Hz	0.021 % of reading + 0.1 mV	
	(20 to 40) Hz	82 $\mu$ V/V + 32 $\mu$ V	
	40 Hz to 20 kHz	39 $\mu$ V/V + 22 $\mu$ V	
	(20 to 50) kHz	44 $\mu$ V/V + 62 $\mu$ V	
	(50 to 100) kHz	71 $\mu$ V/V + 55 $\mu$ V	
	(100 to 300) kHz	0.3 mV/V + 0.12 mV	
	(300 to 500) kHz	0.89 mV/V + 0.3 mV	
	500 kHz to 1 MHz	1.5 mV/V + 0.6 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.2 mV/V + 1.1 mV	
	(20 to 40) Hz	68 $\mu$ V/V + 0.63 mV	
	40 Hz to 20 kHz	19 $\mu$ V/V + 0.7 mV	
	(20 to 50) kHz	44 $\mu$ V/V + 0.61 mV	
	(50 to 100) kHz	0.006 1 % of reading + 0.65 mV	
(100 to 300) kHz	0.19 mV/V + 1.9 mV		
(300 to 500) kHz	0.9 mV/V + 2.8 mV		
500 kHz to 1 MHz	1.1 mV/V + 11 mV		
(22 to 220) V			
(10 to 20) Hz	0.2 mV/V + 11 mV		
(20 to 40) Hz	44 $\mu$ V/V + 17 mV		
40 Hz to 20 kHz	38 $\mu$ V/V + 4.1 mV		
(20 to 50) kHz	53 $\mu$ V/V + 8 mV		
(50 to 100) kHz	0.11 mV/V + 10 mV		
(100 to 300) kHz	0.84 mV/V + 19 mV		
(300 to 500) kHz	4.2 mV/V + 45 mV		
500 kHz to 1 MHz	7.7 mV/V + 90 mV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source <sup>1</sup>	(220 to 1 100) V 40 Hz to 1 kHz (1 to 20) kHz (20 to 30) kHz (220 to 750) V (30 to 50) kHz (50 to 100) kHz	50 $\mu$ V/V + 27 mV 0.1 mV/V + 31 mV 0.41 mV/V + 70 mV 0.45 mV/V + 12 mV 1.8 mV/V + 21 mV	High Performance Multifunction Calibrator w/ Transconductance Amplifier  Vista, CA Orlando, FL
AC Voltage – Source	(1 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz (22 to 220) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.55 mV/V + 4.5 $\mu$ V 0.21 mV/V + 4.5 $\mu$ V 0.11 mV/V + 4.5 $\mu$ V 0.37 mV/V + 4.5 $\mu$ V 0.85 mV/V + 7 $\mu$ V 1.1 mV/V + 13 $\mu$ V 1.7 mV/V + 25 $\mu$ V 3.4 mV/V + 25 $\mu$ V 0.55 mV/V + 5 $\mu$ V 0.21 mV/V + 5 $\mu$ V 0.11 mV/V + 5 $\mu$ V 0.37 mV/V + 5 $\mu$ V 0.85 mV/V + 7 $\mu$ V 1.1 mV/V + 12 $\mu$ V 1.7 mV/V + 25 $\mu$ V 3.4 mV/V + 25 $\mu$ V 0.55 mV/V + 13 $\mu$ V 0.21 mV/V + 8 $\mu$ V 0.11 mV/V + 8 $\mu$ V 0.32 mV/V + 8 $\mu$ V 0.85 mV/V + 25 $\mu$ V 1.1 mV/V + 25 $\mu$ V 1.7 mV/V + 35 $\mu$ V 3.4 mV/V + 80 $\mu$ V	High Performance Multifunction Calibrator w/ Transconductance Amplifier  Fremont, CA

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source	220 mV to 2.2 V		High Performance Multifunction Calibrator w/ Transconductance Amplifier  Fremont, CA
	(10 to 20) Hz	0.24 mV/V + 45 $\mu$ V	
	(20 to 40) Hz	90 $\mu$ V/V + 25 $\mu$ V	
	40 Hz to 20 kHz	45 $\mu$ V/V + 21 $\mu$ V	
	(20 to 50) kHz	75 $\mu$ V/V + 22 $\mu$ V	
	(50 to 100) kHz	0.11 mV/V + 82 $\mu$ V	
	(100 to 300) kHz	0.42 mV/V + 0.11 mV	
	(300 to 500) kHz	1 mV/V + 0.22 mV	
	500 kHz to 1 MHz	1.7 mV/V + 0.31 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.24 mV/V + 0.44 mV	
	(20 to 40) Hz	90 $\mu$ V/V + 0.24 mV	
	40 Hz to 20 kHz	45 $\mu$ V/V + 0.19 mV	
	(20 to 50) kHz	75 $\mu$ V/V + 0.11 mV	
	(50 to 100) kHz	0.1 mV/V + 0.59 mV	
	(100 to 300) kHz	0.28 mV/V + 0.81 mV	
	(300 to 500) kHz	1 mV/V + 2.1 V	
	500 kHz to 1 MHz	1.5 mV/V + 3.3 V	
	(22 to 220) V		
	(10 to 20) Hz	0.24 mV/V + 4.4 mV	
	(20 to 40) Hz	90 $\mu$ V/V + 2.3 mV	
	40 Hz to 20 kHz	52 $\mu$ V/V + 1.8 mV	
	(20 to 50) kHz	80 $\mu$ V/V + 1.9 mV	
	(50 to 100) kHz	0.15 mV/V + 5.3 mV	
	(100 to 300) kHz	0.9 mV/V + 17 mV	
	(300 to 500) kHz	4.4 mV/V + 41 mV	
	500 kHz to 1 MHz	8 mV/V + 81 mV	
	(220 to 250) V		
	(15 to 50) Hz	0.3 mV/V + 17 mV	
	50 Hz to 1 kHz	70 $\mu$ V/V + 3.9 mV	
(1 to 20) kHz	0.17 mV/V + 6.8 mV		
(20 to 30) kHz	0.6 mV/V + 12 mV		
220 V to 1.1 kV			
40 Hz to 1 kHz	90 $\mu$ V/V + 5.3 mV		
(1 to 20) kHz	0.17 mV/V + 6.8 mV		
(20 to 30) kHz	0.6 mV/V + 12 mV		
(220 to 750) V			
(30 to 50) kHz	0.6 mV/V + 12 mV		
(50 to 100) kHz	2.3 mV/V + 46 mV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(1 to 10) mV		Precision 8.5 Digit Multimeter  Fremont, CA Orlando, FL
	(1 to 40) Hz	0.36 mV/V + 3.6 μV	
	40 Hz to 1 kHz	0.23 mV/V + 1.4 μV	
	(1 to 20) kHz	0.32 mV/V + 1.7 μV	
	(20 to 50) kHz	1.2 mV/V + 1.6 μV	
	(50 to 100) kHz	5.7 mV/V + 2 μV	
	(100 to 300) kHz	46 mV/V + 2.6 μV	
	(10 to 100) mV		
	(1 Hz to 40) Hz	72 μV/V + 0.85 mV	
	40 Hz to 1 kHz	84 μV/V + 0.26 mV	
	(1 to 20) kHz	0.17 mV/V + 0.26 mV	
	100 mV to 10 V		
	(1 Hz to 40) Hz	71 μV/V + 0.73 mV	
	40 Hz to 1 kHz	80 μV/V + 0.26 mV	
	(1 to 20) kHz	160 μV/V + 0.26 mV	
	(20 to 50) kHz	0.35 mV/V + 0.25 mV	
	(50 to 100) kHz	0.93 mV/V + 0.27 mV	
	(100 to 300) kHz	3.5 mV/V + 1.2 mV	
	300 kHz to 1 MHz	12 mV/V + 1.2 mV	
	(1 to 2) MHz	18 mV/V + 1.2 mV	
(10 to 100) V			
(1 to 40) Hz	0.24 mV/V + 4.7 mV		
40 Hz to 1 kHz	0.15 mV/V + 16 mV		
(1 to 20) kHz	0.15 mV/V + 17 mV		
(20 to 50) kHz	0.36 mV/V + 8.6 mV		
(50 to 100) kHz	1.4 mV/V + 3.9 mV		
(100 to 300) kHz	4.8 mV/V + 12 mV		
300 kHz to 1 MHz	18 mV/V + 12 mV		
(100 to 1 000) V			
(1 to 40) Hz	0.46 mV/V + 50 mV		
40 Hz to 1 kHz	0.46 mV/V + 28 mV		
(1 to 20) kHz	0.69 mV/V + 29 mV		
(20 to 50) kHz	1.5 mV/V + 24 mV		
(50 to 100) kHz	3.6 mV/V + 24 mV		
AC High Voltage - Measure <sup>1</sup>	(Up to 10) kV		High Voltage Meter  Vista, CA Fremont, CA Orlando, FL
	(30 to 200) Hz	1.4 mV/V + 0.14 V	
	(200 to 450) Hz	4.6 mV/V + 0.14V	
	(450 to 600) Hz	8.7 mV/V + 0.14V	
	(10 to 100) kV		
	(30 to 70) Hz	1.4 mV/V + 0.7 V	
	(70 to 200) Hz	17 % of reading + 0.7 V	

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	Up to 200 mV		Precision 8.5 Digit Multimeter  Vista, CA
	(1 to 10) Hz	0.2 mV/V + 29 μV	
	(10 to 40) Hz	0.17 mV/V + 5 μV	
	(40 to 100) Hz	0.14 mV/V + 5 μV	
	100 Hz to 2 kHz	0.17 mV/V + 2 μV	
	(2 to 10) kHz	0.17 mV/V + 5 μV	
	(10 to 30) kHz	0.42 mV/V + 10 μV	
	(30 to 100) kHz	0.93 mV/V + 24 μV	
	200 mV to 2 V		
	(1 to 10) Hz	0.19 mV/V + 0.33 mV	
	(10 to 40) Hz	0.15 mV/V + 3 μV	
	(40 to 100) Hz	0.12 mV/V + 24 μV	
	100 Hz to 2 kHz	0.17 mV/V + 24 μV	
	(2 to 10) kHz	0.14 mV/V + 24 μV	
	(10 to 30) kHz	0.27 mV/V + 48 μV	
	(30 to 100) kHz	0.7 mV/V + 0.24 mV	
	(2 to 20) V		
	(1 to 10) Hz	0.18 mV/V + 3.3 mV	
	(10 to 40) Hz	0.14 mV/V + 0.3 mV	
	(40 to 100) Hz	0.11 mV/V + 0.24 mV	
	100 Hz to 2 kHz	0.15 mV/V + 0.24 mV	
	(2 to 10) kHz	0.14 mV/V + 0.24 mV	
	(10 to 30) kHz	0.27 mV/V + 0.48 mV	
	(30 to 100) kHz	0.69 mV/V + 2.4 mV	
	(100 to 300) kHz	3.6 mV/V + 24 mV	
	300 kHz to 1 MHz	12 mV/V + 0.24 mV	
	(20 to 200) V		
	1 to 10 Hz	0.19 mV/V + 58 mV	
(10 to 40) Hz	0.15 mV/V + 2.7 mV		
(40 to 100) Hz	0.12 mV/V + 2.4 mV		
100 Hz to 2 kHz	0.15 mV/V + 2.4 mV		
(2 to 10) kHz	0.14 mV/V + 2.4 mV		
(10 to 30) kHz	0.27 mV/V + 4.8 mV		
(30 to 100) kHz	0.69 mV/V + 24 mV		
(100 to 300) kHz	3.6 mV/V + 0.24 mV		
300 kHz to 1 MHz	12 mV/V + 2.4 V		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	200 V to 1 kV (1 to 10) Hz (10 to 40) Hz 40 Hz to 10 kHz (10 to 30) kHz (30 to 100) kHz	0.19 mV/V + 84 mV 0.16 mV/V + 24 mV 0.29 mV/V + 24 mV 0.28 mV/V + 48 mV 0.72 mV/V + 0.24 V	Precision 8.5 Digit Multimeter  Vista, CA
AC Voltage Flatness - Source (Referenced to 1 kHz)	0.3 mV to 3.5 V (10 to 30) Hz 30 Hz to 120 kHz (0.3 to 1.1) mV 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz (1.1 to 3) mV 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz 3 mV to 3.5 V 120 kHz to 2 MHz (2 to 10) MHz (10 to 20) MHz (20 to 30) MHz	0.35 % of reading 0.12 % of reading 0.24 % of reading + 3 μV 0.47 % of reading + 3 μV 0.7 % of reading + 3 μV 1.8 % of reading + 15 μV 0.12 % of reading + 3 μV 0.35 % of reading + 3 μV 0.58 % of reading + 3 μV 1.8 % of reading + 3 μV 0.12 % of reading + 3 μV 0.24 % of reading + 3 μV 0.47 % of reading + 3 μV 1.2 % of reading + 3 μV	High Performance Multifunction Calibrator (Wideband)  Vista, CA Fremont, CA Orlando, FL
AC Voltage Flatness - Measure	Up to 3 V (10 to 100) Hz 100 Hz to 10 kHz (10 to 30) kHz 30 kHz to 1 MHz (1 to 10) MHz (10 to 30) MHz (30 to 50) MHz (50 to 70) MHz (70 to 80) MHz (80 to 100) MHz	0.2 % of reading 0.051 % of reading 0.059 % of reading 0.13 % of reading 0.21 % of reading 0.26 % of reading 0.42 % of reading 0.65 % of reading 0.77 % of reading 0.97 % of reading	Precision 8.5 Digit Multimeter w/ Thermal Converters  Vista, CA Fremont, CA Orlando, FL

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	Up to 2.2 mV		AC Measurement Standard  Vista, CA Orlando, FL
	(10 to 20) Hz	1.6 mV/V + 1.6 μV	
	(20 to 40) Hz	1.6 mV/V + 1.6 μV	
	40 Hz to 20 kHz	0.4 mV/V + 1.7 μV	
	(20 to 50) kHz	0.8 mV/V + 2.2 μV	
	(50 to 100) kHz	1.2 mV/V + 2.7 μV	
	(100 to 300) kHz	2.3 mV/V + 4.1 μV	
	(300 to 500) kHz	2.4 mV/V + 8.1 μV	
	500 kHz to 1 MHz	3.5 mV/V + 8.1 μV	
	(2.2 to 7) mV		
	(10 to 20) Hz	0.8 mV/V + 1.4 μV	
	(20 to 40) Hz	0.4 mV/V + 1.5 μV	
	40 Hz to 20 kHz	0.2 mV/V + 1.5 μV	
	(20 to 50) kHz	0.4 mV/V + 2.1 μV	
	(50 to 100) kHz	0.6 mV/V + 2.6 μV	
	(100 to 300) kHz	1.2 mV/V + 4.1 μV	
	(300 to 500) kHz	1.3 mV/V + 6 μV	
	500 kHz to 1 MHz	2 mV/V + 6 μV	
	(7 to 22) mV		
	(10 to 20) Hz	0.3 mV/V + 1.6 μV	
	(20 to 40) Hz	0.3 mV/V + 1.6 μV	
	40 Hz to 20 kHz	0.1 mV/V + 1.7 μV	
	(20 to 50) kHz	0.2 mV/V + 2.3 μV	
	(50 to 100) kHz	0.3 mV/V + 2.7 μV	
	(100 to 300) kHz	0.8 mV/V + 4.1 μV	
	(300 to 500) kHz	0.9 mV/V + 6.1 μV	
	500 kHz to 1 MHz	1.4 mV/V + 6.1 μV	
	(22 to 70) mV		
(10 to 20) Hz	2 mV/V + 5.3 μV		
(20 to 40) Hz	80 μV/V + 6.3 μV		
40 Hz to 20 kHz	40 μV/V + 6.8 μV		
(20 to 50) kHz	0.1 mV/V + 6.4 μV		
(50 to 100) kHz	0.23 mV/V + 5.8 μV		
(100 to 300) kHz	0.5 mV/V + 6 μV		
(300 to 500) kHz	0.7 mV/V + 9.4 μV		
500 kHz to 1 MHz	1.1 mV/V + 9 μV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(70 to 220) mV		AC Measurement Standard  Vista, CA Orlando, FL
	(10 to 20) Hz	0.2 mV/V + 1.6 μV	
	(20 to 40) Hz	80 μV/V + 1.7 μV	
	40 Hz to 20 kHz	40 μV/V + 1.8 μV	
	(20 to 50) kHz	70 μV/V + 2.2 μV	
	(50 to 100) kHz	0.2 mV/V + 2.6 μV	
	(100 to 300) kHz	0.3 mV/V + 4.1 μV	
	(300 to 500) kHz	0.4 mV/V + 8 μV	
	500 kHz to 1 MHz	1 mV/V + 8 μV	
	(220 to 700) mV		
	(10 to 20) Hz	0.21 mV/V + 1.5 μV	
	(20 to 40) Hz	0.21 mV/V + 1.5 μV	
	40 Hz to 20 kHz	30 μV/V + 1.7 μV	
	(20 to 50) kHz	50 μV/V + 2.1 μV	
	(50 to 100) kHz	80 μV/V + 2.6 μV	
	(100 to 300) kHz	0.2 mV/V + 4 μV	
	(300 to 500) kHz	0.3 mV/V + 8 μV	
	500 kHz to 1 MHz	0.3 mV/V + 8 μV	
	700 mV to 2.2 V		
	(10 to 20) Hz	0.2 mV/V + 4.7 μV	
	(20 to 40) Hz	0.62 mV/V + 13 μV	
	40 Hz to 20 kHz	0.17 mV/V + 23 μV	
	(20 to 50) kHz	0.41 mV/V + 16 μV	
	(50 to 100) kHz	0.67 mV/V + 12 μV	
	(100 to 300) kHz	0.16 mV/V + 6 μV	
	(300 to 500) kHz	0.26 mV/V + 4 μV	
	500 kHz to 1 MHz	9 mV/V + 1.1 μV	
	(2.2 to 7) V		
(10 to 20) Hz	0.2 mV/V + 1.1 μV		
(20 to 40) Hz	0.7 mV/V + 3.2 μV		
40 Hz to 20 kHz	0.23 mV/V + 8.4 μV		
(20 to 50) kHz	0.5 mV/V + 4.4 μV		
(50 to 100) kHz	0.8 mV/V + 2.7 μV		
(100 to 300) kHz	0.19 mV/V + 1.1 μV		
(300 to 500) kHz	0.4 mV/V + 0.5 μV		
500 kHz to 1 MHz	1.2 mV/V + 0.2 μV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure <sup>1</sup>	(7 to 22) V		AC Measurement Standard  Vista, CA Orlando, FL
	(10 to 20) Hz	0.2 mV/V + 3.7 μV	
	(20 to 40) Hz	0.7 mV/V + 11 μV	
	40 Hz to 20 kHz	0.3 mV/V + 26 μV	
	(20 to 50) kHz	0.5 mV/V + 15 μV	
	(50 to 100) kHz	0.8 mV/V + 9.1 μV	
	(100 to 300) kHz	0.29 mV/V + 4 μV	
	(300 to 500) kHz	0.4 mV/V + 1.9 μV	
	500 kHz to 1 MHz	1.2 mV/V + 0.6 μV	
	(22 to 70) V		
	(10 to 20) Hz	0.2 mV/V + 9.1 μV	
	(20 to 40) Hz	70 μV/V + 27 μV	
	40 Hz to 20 kHz	30 μV/V + 55 μV	
	(20 to 50) kHz	60 μV/V + 32 μV	
	(50 to 100) kHz	90 μV/V + 19 μV	
	(100 to 300) kHz	0.2 mV/V + 9.1 μV	
	(300 to 500) kHz	0.4 mV/V + 4.4 μV	
	500 kHz to 1 MHz	1.2 mV/V + 1.5 μV	
	(70 to 220) V		
	(10 to 20) Hz	0.20 mV/V + 7.1 μV	
	(20 to 40) Hz	70 μV/V + 21 μV	
	40 Hz to 20 kHz	30 μV/V + 45 μV	
	(20 to 50) kHz	70 μV/V + 21 μV	
	(50 to 100) kHz	0.1 mV/V + 14 μV	
	(100 to 300) kHz	0.21 mV/V + 6.8 μV	
	(300 to 500) kHz	0.5 mV/V + 2.8 μV	
	(220 to 700) V		
	(10 to 20) Hz	0.2 mV/V + 77 μV	
	(20 to 40) Hz	0.1 mV/V + 0.16 mV	
	40 Hz to 20 kHz	40 μV/V + 0.37 mV	
(20 to 50) kHz	0.13 mV/V + 0.12 mV		
(50 to 100) kHz	0.5 mV/V + 31 μV		
700 V to 1 kV			
(10 to 20) Hz	0.2 mV/V + 31 μV		
(20 to 40) Hz	0.1 mV/V + 62 μV		
40 Hz to 20 kHz	40 μV/V + 0.16 mV		
(20 to 50) kHz	0.13 mV/V + 47 μV		
(50 to 100) kHz	0.5 mV/V + 12 μV		



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage - Measure	Up to 10 mV		Precision 8.5 Digit Multimeter ACBAND < 2 MHz  Vista, CA Fremont, CA Orlando, FL
	(1 to 40) Hz	3.5 $\mu$ V/V + 4.6 $\mu$ V	
	40 Hz to 1 kHz	2.4 $\mu$ V/V + 3.2 $\mu$ V	
	(1 to 20) kHz	3.5 $\mu$ V/V + 4.6 $\mu$ V	
	(20 to 50) kHz	12 $\mu$ V/V + 3.2 $\mu$ V	
	(50 to 100) kHz	58 $\mu$ V/V + 3.2 $\mu$ V	
	(100 to 300) kHz	0.47 mV/V + 3.8 $\mu$ V	
	(10 to 100) mV		
	(1 to 40) Hz	86 $\mu$ V/V + 4.6 $\mu$ V	
	40 Hz to 1 kHz	86 $\mu$ V/V + 2.3 $\mu$ V	
	(1 to 20) kHz	0.17 mV/V + 2.3 $\mu$ V	
	(20 to 50) kHz	0.35 mV/V + 2.3 $\mu$ V	
	(50 to 100) kHz	0.93 mV/V + 2.3 $\mu$ V	
	(100 to 300) kHz	3.5 mV/V + 12 $\mu$ V	
	300 kHz to 1 MHz	12 mV/V + 12 $\mu$ V	
	(1 to 2) MHz	18 mV/V + 12 $\mu$ V	
	100 mV to 1 V		
	(1 to 40) Hz	92 $\mu$ V/V + 50 $\mu$ V	
	40 Hz to 1 kHz	92 $\mu$ V/V + 31 $\mu$ V	
	(1 to 10) V		
	(1 to 20) kHz	0.17 mV/V + 31 $\mu$ V	
	(20 to 50) kHz	0.35 mV/V + 31 $\mu$ V	
	(50 to 100) kHz	0.93 mV/V + 31 $\mu$ V	
	(100 to 300) kHz	3.5 mV/V + 0.12 mV	
300 kHz to 1 MHz	12 mV/V + 0.12 mV		
(1 to 2) MHz	18 mV/V + 0.12 mV		
(10 to 100) V			
(1 to 40) Hz	91 $\mu$ V/V + 0.14 mV		
40 Hz to 1 kHz	91 $\mu$ V/V + 0.14 mV		
(1 to 20) kHz	0.17 mV/V + 0.14 mV		
(20 to 50) kHz	0.35 mV/V + 0.14 mV		
(50 to 100) kHz	0.93 mV/V + 0.14 mV		
(100 to 300) kHz	3.5 mV/V + 0.16 mV		
300 kHz to 1 MHz	1.2 mV/V + 0.16 mV		
(1 to 2) MHz	1.8 mV/V + 0.16 mV		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage - Measure	(100 to 750) V (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz 300 kHz to 1 MHz	0.45 mV/V + 2.7 mV 0.45 mV/V + 2.7 mV 0.45 mV/V + 2.7 mV 0.56 mV/V + 2.7 mV 1.5 mV/V + 2.7 mV 4.7 mV/V + 2.7 mV 18 mV/V + 2.7 mV	Precision 8.5 Digit Multimeter ACBAND < 2 MHz  Vista, CA Fremont, CA Orlando, FL
AC Voltage - Measure	Up to 10 mV (1 to 40) Hz 40 Hz to 1 kHz (1 to 20) kHz (20 to 50) kHz (50 to 100) kHz 11 mV to 10 V 45 Hz to 100 kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (11 to 100) V 45 Hz to 100 kHz 100 kHz to 1 MHz (1 to 4) MHz (4 to 8) MHz (8 to 10) MHz (100 to 750) V 45 Hz to 100 kHz	0.46 mV/V + 47 mV 0.46 mV/V + 24 mV 0.69 mV/V + 24 mV 1.4 mV/V + 24 mV 3.5 mV/V + 24 mV 1 mV/V + 7.5 μV 14 mV/V + 6.5 μV 81 mV/V + 8.6 μV 0.24 V + 9.7 μV 1 mV/V + 0.7 mV 24 mV/V + 0.58 mV 47 mV/V + 0.81 mV 47 mV/V + 0.93 mV 0.18 V + 1.2 mV 1.5 mV/V + 2.7 mV	Precision 8.5 Digit Multimeter ACBAND > 2 MHz  Vista, CA Fremont, CA Orlando, FL
AC Current – Source <sup>1</sup>	(9 to 220) μA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 μA to 2.2 mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.02 % of reading + 26 nA 0.01 % of reading + 30 nA 0.01 % of reading + 8.3 nA 0.02 % of reading + 36 nA 0.1 % of reading + 63 nA  0.004 8 % of reading + 1.4 μA 0.006 1 % of reading + 0.44 μA 0.007 2 % of reading + 95 nA 0.016 % of reading + 0.18 μA 0.1 % of reading + 0.65 μA	High Performance Multifunction Calibrator  Vista, CA Orlando, FL



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source <sup>1</sup>	(2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (22 to 220 mA) (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz 220 mA to 2.2 A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.003 4 % of reading + 20 μA 0.002 5 % of reading + 12 μA 0.007 2 % of reading + 0.95 μA 0.015 % of reading + 1.5 μA 0.1 % of reading + 5.1 μA 0.004 6 % of reading + 0.15 mA 0.003 5 % of reading + 85 μA 0.006 2 % of reading + 12 μA 0.014 % of reading + 16 μA 0.089 % of reading + 38 μA 0.18 mA/A + 0.17 mA 0.32 mA/A + 0.27 mA 5.1 mA/A + 2.8 mA 0.3 mA/A + 1.1 mA 0.74 mA/A + 53 μA 2.8 mA/A + 200 μA	High Performance Multifunction Calibrator  Vista, CA Orlando, FL
AC Current – Source <sup>1</sup>	(11 to 20) A (45 to 100) Hz 100 Hz to 1kHz (1 to 5) kHz	1.2 mA/A + 5 mA 1.5 mA/A + 5 mA 30 mA/A + 5 mA	Multi Product Calibrator  Vista, CA Orlando, FL
AC Current – Source <sup>1</sup> Clamp-On Meters	(10 to 16.5) A (45 to 65) Hz (65 to 440) Hz (16.5 to 150) A (45 to 65) Hz (65 to 440) Hz (150 to 1 025) A (45 to 65) Hz (65 to 440) Hz	0.28 % of reading + 0.16 A 0.79 % of reading + 0.13 A 0.28 % of reading + 0.31 A 0.79 % of reading + 0.16 A 0.29 % of reading + 1.3 A 0.8 % of reading + 1.2 A	Multi Product Calibrator w/ Current Coil  Vista, CA Fremont, CA Orlando, FL

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source	Up to 220 $\mu$ A		High Performance Multifunction Calibrator  Fremont, CA
	(10 to 20) Hz	0.25 mA/A + 16 nA	
	(20 to 40) Hz	0.16 mA/A + 10 nA	
	40 Hz to 1 kHz	0.12 mA/A + 8 nA	
	(1 to 5) kHz	0.28 mA/A + 12 nA	
	(5 to 10) kHz	1.1 mA/A + 65 nA	
	220 $\mu$ A to 2.2 mA		
	(10 to 20) Hz	0.25 mA/A + 40 nA	
	(20 to 40) Hz	0.16 mA/A + 36 nA	
	40 Hz to 1 kHz	0.12 mA/A + 36 nA	
	(1 to 5) kHz	0.2 mA/A + 0.11 $\mu$ A	
	(5 to 10) kHz	1.1 mA/A + 0.65 $\mu$ A	
	(2.2 to 22) mA		
	(10 to 20) Hz	0.25 mA/A + 0.41 $\mu$ A	
	(20 to 40) Hz	0.16 mA/A + 0.36 $\mu$ A	
	40 Hz to 1 kHz	0.12 mA/A + 0.36 $\mu$ A	
	(1 to 5) kHz	0.2 mA/A + 0.56 $\mu$ A	
	(5 to 10) kHz	1.1 mA/A + 5.1 $\mu$ A	
	(22 to 220) mA		
	(10 to 20) Hz	0.25 mA/A + 4.1 $\mu$ A	
	(20 to 40) Hz	0.16 mA/A + 3.6 $\mu$ A	
	40 Hz to 1 kHz	0.12 mA/A + 2.6 $\mu$ A	
	(1 to 5) kHz	0.2 mA/A + 3.6 $\mu$ A	
	(5 to 10) kHz	1.1 mA/A + 11 $\mu$ A	
220 mA to 2.2 A			
20 Hz to 1 kHz	0.25 mA/A + 36 $\mu$ A		
(1 to 5) kHz	0.45 mA/A + 81 $\mu$ A		
(5 to 10) kHz	7 mA/A + 0.18 mA		
(2.2 to 11) A			
40 Hz to 1 kHz	0.46 mA/A + 0.21 mA		
(1 to 5) kHz	0.95 mA/A + 0.4 mA		
(5 to 10) kHz	3.6 mA/A + 0.76 mA		
(11 to 20) A			
(45 to 100) Hz	1.2 mA/A + 5 mA		
100 Hz to 1kHz	1.5 mA/A + 5 mA		
(1 to 5) kHz	30 mA/A + 5 mA		

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source <sup>1</sup> PF = 1 (10 to 45) Hz 33 mV to 32.9999 V (3.3 mA to 2.99999 A)  (45 to 65) Hz 33 mV to 1020 V (3.3 mA to 20.5 A)	110 μW to 99 W  110 μW to 21 kW	0.19 % of reading  0.15 % of reading	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
AC Power - Measure	Up to 15 W (0.1 to 66) Hz 66 Hz to 1 kHz (>15 to 30) W (0.1 to 66) Hz 66 Hz to 1 kHz (>30 to 60) W (0.1 to 66) Hz 66 Hz to 1 kHz (>60 to 100) W (0.1 to 66) Hz 66 Hz to 1 kHz (>100 to 150) W (0.1 to 66) Hz 66 Hz to 1 kHz (>150 to 300) W (0.1 to 66) Hz 66 Hz to 1 kHz (>300 to 600) W (0.1 to 66) Hz 66 Hz to 1 kHz	0.1 % of reading + 0.012 W 0.068 % of reading + 0.018 W  0.1 % of reading + 0.023 W 0.068 % of reading + 0.035 W  0.1 % of reading + 0.046 W 0.068 % of reading + 0.071 W  0.1 % of reading + 0.076 W 0.068 % of reading + 0.12 W  0.1 % of reading + 0.12 W 0.068 % of reading + 0.18 W  0.1 % of reading + 0.24 W 0.068 % of reading + 0.36 W  0.1 % of reading + 0.43 W 0.068 % of reading + 0.71 W	Precision Power Analyzer  Vista, CA
AC Power - Measure	(>600 to 1 000) W (0.1 to 66) Hz 66 Hz to 1 kHz (>1 000 to 2 000) W (0.1 to 66) Hz 66 Hz to 1 kHz	0.1 % of reading + 0.45 W 0.068 % of reading + 1.2 W  0.1 % of reading + 0.81 W 0.068 % of reading + 2.4 W	Precision Power Analyzer  Vista, CA



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Phase – Source <sup>1</sup> (10 to 65) Hz (65 to 500) Hz 500 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	+/- 179.99 ( $\Delta\Phi^\circ$ )	0.09° 0.2° 0.39° 1.9° 3.9° 7.8°	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
Phase – Measure	(0 to 360)° 10 Hz to 50 kHz (50 to 100) kHz	0.084° 0.41°	Phase Meter  Vista, CA
AC Current – Measure <sup>1</sup>	Up to 100 $\mu$ A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 1 kHz 100 $\mu$ A to 100 mA (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz 100 $\mu$ A to 100 mA (20 to 50) kHz (50 to 100) kHz 100 mA to 1 A (10 to 20) Hz (20 to 45) Hz (45 to 100) Hz 100 Hz to 5 kHz (5 to 20) kHz (20 to 50) kHz	4.7 mA/A + 36 nA 1.8 mA/A + 35 nA 0.71 mA/A + 35 nA 0.7 mA/A + 35 nA 4.7 mA/A + 24 $\mu$ A 1.8 mA/A + 24 $\mu$ A 0.71 mA/A + 24 $\mu$ A 0.36 mA/A + 23 $\mu$ A 0.71 mA/A + 24 $\mu$ A 4.7 mA/A + 47 $\mu$ A 6.5 mA/A + 0.18 mA 4.7 mA/A + 0.24 mA 1.9 mA/A + 0.24 mA 0.95 mA/A + 0.24 mA 1.1 mA/A + 0.26 mA 3.5 mA/A + 0.24 mA 12 mA/A + 0.47 mA	Precision 8.5 Digit Multimeter  Fremont, CA Orlando, FL
AC Current – Measure <sup>1</sup>	(1 to 10) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz (5 to 10) kHz (10 to 100) A (50 to 60) Hz (100 to 300) A (50 to 60) Hz	20 mA/A + 6 mA 11 mA/A + 6 mA 1.5 mA/A + 6 mA 3.5 mA/A + 70 mA  0.6 A  0.8 A	Precision 8.5 Digit Multimeter / Shunts  Vista, CA Fremont, CA Orlando, FL

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current - Measure	Up to 200 $\mu$ A		Precision 8.5 Digit Multimeter  Vista CA
	10 Hz to 10 kHz	0.42 mA/A + 24 nA	
	(10 to 30) kHz	0.88 mA/A + 24 nA	
	(30 to 100) kHz	4.8 mA/A + 24 nA	
	200 $\mu$ A to 2 mA		
	10 Hz to 10 kHz	0.36 mA/A + 0.24 $\mu$ A	
	(10 to 30) kHz	0.86 mA/A + 0.24 $\mu$ A	
	(30 to 100) kHz	4.8 mA/A + 0.24 $\mu$ A	
	(2 to 20) mA		
	10 Hz to 10 kHz	0.37 mA/A + 2.4 $\mu$ A	
	(10 to 30) kHz	0.86 mA/A + 2.4 $\mu$ A	
	(30 to 100) kHz	4.8 mA/A + 2.4 $\mu$ A	
	(20 to 200) mA		
	10 Hz to 10 kHz	0.35 mA/A + 24 $\mu$ A	
(10 to 30) kHz	0.75 mA/A + 24 $\mu$ A		
200 mA to 2 A			
10 Hz to 2 kHz	0.75 mA/A + 0.25 mA		
(2 to 10) kHz	0.87 mA/A + 0.27 mA		
(10 to 30) kHz	3.6 mA/A + 0.26 mA		
(2 to 20) A			
10 Hz to 2 kHz	0.99 mA/A + 2.4 mA		
(2 to 10) kHz	3.1 mA/A + 2.4 mA		
AC High Current - Measure	(50 to 400) H z		Precision 8.5 Digit Multimeter and Current Transducers  Orlando, FL
	(0.25 to 10) A	0.59 % of reading + 0.029 A	
	(10 to 100) A	0.59 % of reading + 0.37 A	
	(100 to 400) A	0.6 % of reading + 1.6 A	
	(400 to 600) A	0.6 % of reading + 2 A	
	(50 to 60) Hz		
	(600 to 900) A	0.6 % of reading + 3.8 A	
(900 to 1 100) A	0.6 % of reading + 4.1 A		
(1 100 to 1 500) A	0.6 % of reading + 4.9 A		
Resistance – Source <sup>1</sup>	Up to 11 $\Omega$	1.7 m $\Omega$ / $\Omega$ + 35 $\mu\Omega$	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
	(11 to 33) $\Omega$	23 $\mu\Omega$ / $\Omega$ + 3.3 m $\Omega$	
	(33 to 110) $\Omega$	31 $\mu\Omega$ / $\Omega$ + 1.9 m $\Omega$	
	(110 to 330) $\Omega$	32 $\mu\Omega$ / $\Omega$ + 2.8 m $\Omega$	
	330 $\Omega$ to 1.1 k $\Omega$	33 $\mu\Omega$ / $\Omega$ + 2.3 m $\Omega$	
	(1.1 to 3.3) k $\Omega$	32 $\mu\Omega$ / $\Omega$ + 30 m $\Omega$	
	(3.3 to 11) k $\Omega$	5 $\mu\Omega$ / $\Omega$ + 1.4 $\Omega$	
	(11 to 33) k $\Omega$	32 $\mu\Omega$ / $\Omega$ + 0.3 $\Omega$	
	(33 to 110) k $\Omega$	33 $\mu\Omega$ / $\Omega$ + 0.34 $\Omega$	

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source <sup>1</sup>	(110 to 330) kΩ 330 kΩ to 1.1 MΩ (1.1 to 3.3) MΩ (3.3 to 11) MΩ (11 to 33) MΩ (33 to 110) MΩ (110 to 330) MΩ 330 MΩ to 1.1 GΩ	37 μΩ/Ω + 3.1 Ω 38 μΩ/Ω + 2.6 Ω 71 μΩ/Ω + 36 Ω 0.13 mΩ/Ω + 0.38 kΩ 0.22 mΩ/Ω + 7.4 kΩ 0.42 mΩ/Ω + 32 kΩ 3.6 mΩ/Ω + 0.12 MΩ 18 mΩ/Ω + 0.59 MΩ	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
Resistance – Source <sup>1</sup> Fixed Points	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	48 μΩ 0.15 mΩ 0.22 mΩ 0.28 mΩ 0.63 mΩ 1.2 mΩ 2.3 mΩ 7.8 mΩ 15 mΩ 77 mΩ 0.15 Ω 1 Ω 1.9 Ω 16 Ω 40 Ω 0.48 kΩ 1.1 kΩ 12 kΩ	High Performance Multifunction Calibrator  Vista, CA Orlando, FL
Resistance – Source <sup>1</sup> Fixed Points	1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 kΩ 1.9 kΩ 10 kΩ 19 kΩ 100 kΩ 190 kΩ 1 MΩ 1.9 MΩ	0.12 mΩ 0.22 mΩ 0.28 mΩ 0.58 mΩ 1.3 mΩ 2.4 mΩ 10 mΩ 20 mΩ 90 mΩ 0.19 Ω 1.3 Ω 2.5 Ω 24 Ω 47 Ω	High Performance Multifunction Calibrator  Fremont, CA

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source <sup>1</sup> Fixed Points	1.9 MΩ 10 MΩ 19 MΩ 100 MΩ	47 Ω 0.47 kΩ 1.3 kΩ 12 kΩ	High Performance Multifunction Calibrator  Fremont, CA
Resistance – Source <sup>1</sup> Fixed Point (Four-Terminal Pair)	0.1 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ	0.56 mΩ 0.29 mΩ 2.3 mΩ 1.8 mΩ 35 mΩ 0.12 Ω 1.3 Ω	Four Terminal Pair Resistor Set  Vista, CA Fremont, CA Orlando, FL
Resistance – Measure <sup>1</sup>	Up to 10 Ω (>10 to 100) Ω >100 Ω to 1 kΩ (>1 to 10) kΩ (>10 to 100) kΩ >100 kΩ to 1 MΩ (>1 to 10) MΩ (>10 to 100) MΩ >100 MΩ to 1 GΩ	21 μΩ/Ω + 62 μΩ 17 μΩ/Ω + 0.59 mΩ 15 μΩ/Ω + 0.67 mΩ 15 μΩ/Ω + 6.1 mΩ 15 μΩ/Ω + 62 mΩ 21 μΩ/Ω + 2.4 Ω 45 μΩ/Ω + 0.38 kΩ 0.58 mΩ/Ω + 1.2 kΩ 5.8 mΩ/Ω + 11 kΩ	Precision 8.5 Digit Multimeter  Fremont, CA Orlando, FL
Resistance – Measure <sup>1</sup>	Up to 2Ω (2 to 20) Ω (20 to 200) Ω 200 Ω to 2 kΩ (2 to 20) kΩ (20 to 200) kΩ 200 kΩ to 2 MΩ (2 to 20) MΩ (20 to 200) MΩ 200 MΩ to 2 GΩ	23 μΩ/Ω + 5.9 μΩ 12 μΩ/Ω + 18 μΩ 9.8 μΩ/Ω + 80 μΩ 12 μΩ/Ω + 0.94 mΩ 11 μΩ/Ω + 47 mΩ 11 μΩ/Ω + 60 mΩ 13 μΩ/Ω + 1.2 Ω 28 μΩ/Ω + 0.12 kΩ 0.15 mΩ/Ω + 1.2 kΩ 1.9 mΩ/Ω + 12 kΩ	Precision 8.5 Digit Multimeter  Vista, CA
Capacitance – Source <sup>1</sup> 10 Hz to 10 kHz 10 Hz to 10 kHz 10 Hz to 3 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz 10 Hz to 1 kHz (10 to 600) Hz (10 to 300) Hz	(220 to 399.9) pF (0.4 to 1.099 9) nF (1.1 to 3.299 9) nF (3.3 to 10.999 9) nF (11 to 32.999 9) nF (33 to 109.999) nF (110 to 329.999) nF (0.33 to 1.099 99) μF (1.1 to 3.299 99) μF	0.58 % of reading + 12 pF 0.57 % of reading + 12 pF 0.57 % of reading + 12 pF 0.22 % of reading + 27 pF 0.29 % of reading + 0.12 nF 0.29 % of reading + 0.13 nF 0.29 % of reading + 0.35 nF 0.28 % of reading + 1.5 nF 0.29 % of reading + 3.5 nF	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL



ANSI National Accreditation Board

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source <sup>1</sup> (10 to 150) Hz (10 to 120) Hz (10 to 80) Hz (0 to 50) Hz (0 to 20) Hz (0 to 6) Hz (0 to 2) Hz (0 to 0.6) Hz (0 to 0.2) Hz	(3.3 to 10.999 9) $\mu$ F (11 to 32.999 9) $\mu$ F (33 to 109.999) $\mu$ F (110 to 329.999) $\mu$ F (0.33 to 1.099 99) mF (1.1 to 3.299 99) mF (3.3 to 10.999 9) mF (11 to 32.999 9) mF (33 to 110) mF	0.29 % of reading + 1.3 nF 0.46 % of reading + 36 nF 0.53 % of reading + 0.12 $\mu$ F 0.53 % of reading + 0.35 $\mu$ F 0.5 % of reading + 1.5 $\mu$ F 0.52 % of reading + 3.6 $\mu$ F 0.51 % of reading + 13 $\mu$ F 0.86 % of reading + 35 $\mu$ F 1.3 % of reading + 0.13 mF	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
Fixed Capacitance @ 1 kHz	(100 to 500) pF 500 pF to 5 nF (5 to 50) nF (50 to 100) nF (100 to 500) nF 500 nF to 1.5 $\mu$ F	0.64 pF 3.7 pF 32 pF 63 pF 0.59 nF 0.63 nF	Capacitance Standard Set  Vista, CA Fremont, CA Orlando, FL
Capacitance – Source <sup>1</sup>	1 pF 1 kHz to 3 MHz 4 MHz 5 MHz 10 MHz 13 MHz 10 pF 1 kHz to 13 MHz 100 pF 1 kHz to 10 MHz 13 MHz 1 nF 1 kHz to 4 MHz 5 MHz 10 MHz 13 MHz	1.2 fF 1.3 fF 1.5 fF 2.8 fF 3.9 fF 12 fF 0.12 pF 0.13 pF 1.2 pF 1.3 pF 2.2 pF 3 pF	Standard Air Capacitor Set  Vista, CA Fremont, CA Orlando, FL
Capacitance - Measure	(1 to 10) nF (10 to 100) nF (0.1 to 1) $\mu$ F (1 to 10) $\mu$ F (10 to 100) $\mu$ F (100 to 1 000) $\mu$ F (1 to 10) mF (10 to 100) mF	0.052 pF 0.041 pF 0.37 nF 0.018 $\mu$ F 0.041 $\mu$ F 0.56 $\mu$ F 0.018 mF 0.059 mF	Digit Multimeter/LCR Meter  Vista, CA Fremont, CA Orlando, FL

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Inductance - Source  0.1 to 1 kHz	200 $\mu$ H 2 mH 20 mH 200 mH 2 H	0.58 $\mu$ H 2.4 $\mu$ H 24 $\mu$ H 0.24 mH 2.4 mH	Standard Value Inductors  Vista, CA Fremont, CA Orlando, FL
Inductance - Measure	0.001 nH to 99.999 999 H Up to 1 kHz 1 kHz to 1 MHz	0.062 % of reading + 0.007 mH 0.56 % of reading + 0.007 mH	Precision LCR Meter  Vista, CA Fremont, CA Orlando, FL
Electrical Simulation of Thermocouple Indicating Devices – Source and Measure <sup>1</sup>	Type C (0 to 150) °C (150 to 650) °C (650 to 1 000) °C (1 000 to 1 800) °C (1 800 to 2 316) °C Type E (-250 to -100) °C (-100 to -25) °C (-25 to 350) °C (350 to 650) °C (650 to 1 000) °C Type J (-210 to -100) °C (-100 to -30) °C (-30 to 150) °C (150 to 760) °C (760 to 1 200) °C Type K (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 1 000) °C (1 000 to 1 372) °C Type N (-200 to -100) °C (-100 to -25) °C (-25 to 120) °C (120 to 410) °C (410 to 1 300) °C	0.29 °C 0.32 °C 0.28 °C 0.48 °C 0.71 °C 0.43 °C 0.21 °C 0.17 °C 0.14 °C 0.19 °C 0.35 °C 0.2 °C 0.14 °C 0.19 °C 0.26 °C 0.28 °C 0.21 °C 0.23 °C 0.25 °C 0.33 °C 0.39 °C 0.27 °C 0.23 °C 0.18 °C 0.27 °C	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
<p>Electrical Simulation of Thermocouple Indicating Devices – Source and Measure <sup>1</sup></p>	<p>Type R            (0 to 250) °C            (250 to 400) °C            (400 to 1 000) °C            (1 000 to 1 767) °C            Type S            (0 to 250) °C            (250 to 1 000) °C            (1 000 to 1 400) °C            (1 400 to 1 767) °C            Type T            (-250 to -150) °C            (-150 to 0) °C            (0 to 120) °C            (120 to 400) °C</p>	<p>0.83 °C            0.42 °C            0.31 °C            0.38 °C            0.55 °C            0.34 °C            0.32 °C            0.41 °C            0.63 °C            0.35 °C            0.26 °C            0.22 °C</p>	<p>Multi Product Calibrator             Vista, CA            Fremont, CA            Orlando, FL</p>
<p>Electrical Simulation of RTD Temperature Indicating Devices – Source and Measure <sup>1</sup></p>	<p>Pt 385 (100 Ω)            (-200 to 0) °C            (0 to 100) °C            (100 to 300) °C            (300 to 400) °C            (400 to 630) °C            (630 to 800) °C            Pt 3926 (100 Ω)            (-200 to 0) °C            (0 to 100) °C            (100 to 300) °C            (300 to 400) °C            (400 to 630) °C            Pt 3916 (100 Ω)            (-200 to -190) °C            (-190 to -80) °C            (-80 to 0) °C            (0 to 100) °C            (100 to 260) °C            (260 to 300) °C            (300 to 400) °C            (400 to 600) °C            (600 to 630) °C</p>	<p>0.05 °C            0.07 °C            0.09 °C            0.1 °C            0.12 °C            0.23 °C            0.05 °C            0.07 °C            0.09 °C            0.1 °C            0.12 °C            0.25 °C            0.04 °C            0.05 °C            0.06 °C            0.07 °C            0.08 °C            0.09 °C            0.1 °C            0.23 °C</p>	<p>Multi Product Calibrator             Vista, CA            Fremont, CA            Orlando, FL</p>

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Temperature Indicating Devices – Source and Measure <sup>1</sup>	Pt 385 (200 Ω)		Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
	(-200 to 100) °C	0.04 °C	
	(100 to 260) °C	0.05 °C	
	(260 to 300) °C	0.12 °C	
	(300 to 400) °C	0.13 °C	
	(400 to 600) °C	0.14 °C	
	(600 to 630) °C	0.16 °C	
	Pt 385 (500 Ω)		
	(-200 to -80) °C	0.04 °C	
	(-80 to 100) °C	0.05 °C	
	(100 to 260) °C	0.06 °C	
	(260 to 400) °C	0.08 °C	
	(400 to 600) °C	0.09 °C	
	(600 to 630) °C	0.11 °C	
	Pt 385 (1 000 Ω)		
	(-200 to 0) °C	0.03 °C	
	(0 to 100) °C	0.04 °C	
(100 to 260) °C	0.05 °C		
(260 to 300) °C	0.06 °C		
(300 to 600) °C	0.07 °C		
(600 to 630) °C	0.23 °C		
PtNi 385 (120 Ω) (Ni 120)			
(-80 to 100) °C	0.08 °C		
(100 to 260) °C	0.14 °C		
CU 427 (10 Ω)			
(100 to 260) °C	0.3 °C		
Oscilloscopes <sup>1</sup> - DC Voltage Into 50 Ω Into 1 MΩ  Square Wave Into 50 Ω 10 Hz to 10 kHz  Into 1 MΩ 10 Hz to 1 kHz (1 to 10) kHz  Leveled Sine Amplitude Reference @ 50 kHz	(0 to +/-6.6) V	0.2 % + 36 μV	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
	(0 to +/-130) V	0.039 % of reading + 37 μV	
	1 mVpp to 6.6 Vpp	0.2 % of reading + 65 μV	
	1 mVpp to 130 Vpp	0.078 % of reading + 36 μV	
1 mVpp to 130 Vpp	0.19 % of reading + 39 μV	15 mV/V + 0.49 mV	
5 mV to 5.5 V			

**Electrical – DC/Low Frequency**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Oscilloscopes <sup>1</sup> - Leveled Sine Amplitude (relative to 50 kHz) 5 mV to 5.5 V  Time Markers (5-2-1 sequence) into 50 Ω  Edge Transition Time (Rise Time)	50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz (600 to 1 100) MHz  (1 to 50) ns 100 ns to 20 ms 50 ms to 5 s  <300 ps	14 mV/V + 0.12 mV 17 mV/V + 0.14 mV 32 mV/V + 0.15 mV 40 mV/V + 0.16 mV  0.001 1 % of reading + 0.048 ps 0.000 2 % of reading + 7 ps 0.4 % of reading  80 ps	Multi Product Calibrator  Vista, CA Fremont, CA Orlando, FL
Oscilloscopes - Square Wave Signal 10 Hz to 10 kHz 50 Ω 1 M Ω DC Signal Into 50Ω Into 1MΩ	1 mVpp to 6.6 V p-p 1 mVpp to 130 V p-p  1 mVpp to 6.6 V p-p 1 mVpp to 130 V p-p	2.5 mV/V + 40 μV 0.5 mV/V + 5 μV  2.5 mV/V + 40 μV 0.5 mV/V + 5 μV	Oscilloscope Calibrator  Vista, CA
Oscilloscopes -  Leveled Sine Wave  Square Wave Signal Leveled Sine Wave Flatness referenced to 50 kHz reference  Time Markers (5-2-1 sequence) into a 50 Ω load  Fast Edge Mode  Programmable Rise Time	Absolute Amplitude 5 mV to 5.5 V 50 kHz Reference 50 kHz to 100 MHz (100 to 300) MHz (300 to 500) MHz (500 to 600) MHz  5 mV to 5.5 V 10 MHz Reference 600 MHz to 1.6 GHz (1.6 to 2.1) GHz  (1 to 50) ns 100 ns to 20 ms 50 ms to 5 s  5 s to 50 ms 20 ms to 500 ps  (1 to 100) kHz	20 mV/V + 0.3 mV 35 mV/V + 0.3 mV 40 mV/V + 0.3 mV 55 mV/V + 0.3 mV 60 mV/V + 0.3 mV  20 mV/V + 0.3 mV 70 mV/V + 0.3 mV 80 mV/V + 0.3 mV  0.001 1 % of reading + 0.048 ps 0.000 2 % of reading + 7 ps 0.4 % of reading  2.5 μs/s + 5 μHz 0.33 μs/s  < 0.15 ns/s +/- 25 ps	Oscilloscope Calibrator  Vista, CA

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Rise Time (measurement)	20 ps to 100 $\mu$ s	81 ns	Oscilloscope  Vista, CA Fremont, CA Orlando, FL
ESD Simulators Contact Voltage (Positive and Negative)	(1 to 8) kV	1.2 % of reading	Multimeters, ESD Target, Attenuator and Oscilloscope
Rise Time	(0.6 to 1) ns	0.000 12 % $\Delta$ time reading + 11 ps	Vista, CA Fremont, CA Orlando, FL
Peak Current	(3.75 to 33) A	2.1 % of reading	
30 ns Current	(2 to 20.8) A	3.4 % of reading	
60 ns Current	(1 to 10.4) A	6.3 % of reading	
ESD Simulators Air Discharge Voltage (Positive and Negative)	(0.7 to 1) ns (1 to 30) kV	1.2 % of reading	Multimeters, ESD Target, Attenuator and Oscilloscope
Rise Time			
RC Time Constant (at $\pm$ 15 kV)	600 ns $\pm$ 130 ns 330 pF probe 300 ns $\pm$ 60 ns 150 pF probe	0.000 12 % $\Delta$ time reading + 11 ps  0.000 12 % $\Delta$ time reading + 11 ps	Vista, CA Fremont, CA Orlando, FL
EFT/Burst Generator Voltage ( $\pm$ )	10 V to 8 kV	2.7 % of reading	Fast Rise Oscilloscope w/ EFT Verification Set
Rise Time	5 ns $\pm$ 30 %	0.008 % $\Delta$ time reading	
Impulse Duration	50 ns $\pm$ 30 %	0.008 % $\Delta$ time reading	Vista, CA Fremont, CA Orlando, FL
Burst Duration	15 ns $\pm$ 20 %	0.008 % $\Delta$ time reading	
Burst Period	300 ms $\pm$ 20 %	0.008 % $\Delta$ time reading	
Surge Generator Front Time			Fast Rise Oscilloscope, Current Probes, Oscilloscope Probes
Rise Time			
Open Circuit ( $\pm$ )	(1.2 to 50) $\mu$ s	0.008 % $\Delta$ time reading	
Short Circuit ( $\pm$ )	(1.2 to 50) $\mu$ s	0.008 % $\Delta$ time reading	
Time to Half-Value ( $\pm$ )	(20 to 700) $\mu$ s	0.008 % $\Delta$ time reading	
Open Circuit Voltage ( $\pm$ )	10 V to 12 kV	2.9 % of reading	Vista, CA Fremont, CA Orlando, FL
Short Circuit Voltage ( $\pm$ )	(0.125 to 3) kA	2.9 % of reading	
Ring Wave Voltage	1 kV $\pm$ 10 %	2.9 % of reading	
Ring Wave Rise Time	1.5 $\mu$ s $\pm$ 0.5 $\mu$ s	0.008 % $\Delta$ time reading	

**Electrical – DC/Low Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Gauss Meter <sup>1</sup>	100 Gauss 200 Gauss 500 Gauss 1 000 Gauss	1.4 Gauss 2.5 Gauss 6.4 Gauss 7.7 Gauss	Reference Magnetic Field Block  Vista, CA

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Attenuation - Source Coaxial, 10 dB Step	(0 to 120) dB 1 kHz 500 MHz 1 GHz	0.11 dB 0.28 dB 0.39 dB	Standard Piston Attenuator  Vista, CA Fremont, CA Orlando, FL
RF Attenuation Source Coaxial, 1 dB Step	DC to 12.4 GHz (0 to 11) dB (12.4 to 18) GHz (0 to 11) dB	0.6 dB 0.8 dB	Standard Electronic Programmable Attenuator  Fremont, CA Orlando, FL Vista, CA
RF Attenuation Source Coaxial, 10 dB Step	DC to 12.4 GHz (0 to 10) dB 20 dB 30 dB 40 dB 50 dB 60 dB 70 dB (80 to 110) dB	0.51 dB 0.71 dB 0.91 dB 1.2 dB 1.5 dB 1.8 dB 2.1 dB 2.4 dB	Standard Electronic Programmable Attenuator  Fremont, CA Orlando, FL Vista, CA
RF Attenuation Source Coaxial, 10 dB Step	(12.4 to 18) GHz (0 to 10) dB 20 dB 30 dB 40 dB 50 dB 60 dB 70 dB (80 to 110) dB	0.61 dB 0.81 dB 1.2 dB 1.6 dB 2 dB 2.4 dB 2.8 dB 3.2 dB	Standard Electronic Programmable Attenuator  Fremont, CA Orlando, FL Vista, CA

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
RF Absolute Power Source Into 50 Ω 0.001 Hz to 100 kHz SWR 1.2:1  100 kHz to 10 MHz SWR ≤ 1.2:1  (10 to 20) MHz SWR ≤ 1.2:1	(3 to 10) V <sub>pp</sub> 1 mV <sub>pp</sub> to 3V <sub>pp</sub>  (3 to 10) V <sub>pp</sub> 1 mV <sub>pp</sub> to 3 V <sub>pp</sub>  100 mV <sub>pp</sub> to 3 V <sub>pp</sub> (1 to 100) mV <sub>pp</sub>	0.13 dB 0.22 dB  0.41 dB 0.61 dB  0.61 dB 0.91 dB	Synthesized Signal Sources  Vista, CA Fremont, CA Orlando, FL
RF Absolute Power Source Into 50 Ω Into 50 Ω/75 Ω	13.01 dBm 1 kHz to 25 MHz 200 Hz to 80 MHz	0.12 dB 0.18 dB	Synthesized Level Generator  Vista, CA
RF Absolute Power Into 50 Ω, In 2 dB steps Relative to Full Output	200 Hz to 80 MHz (0 to -38) dBm (-40 to -58) dBm (-60 to -98) dBm	0.21 dB 0.21 dB 0.27 dB	Synthesized Level Generator  Vista, CA Fremont, CA Orlando, FL
RF Absolute Power Into 75 Ω, In 2 dB steps Relative to Full Output	(0 to -38) dBm 200 Hz to 25 MHz (25 to 80) MHz (-40 to -58) dBm 200 Hz to 25 MHz (25 to 80) MHz (-60 to -98) dBm 200 Hz to 25 MHz (25 to 80) MHz	0.21 dB 0.36 dB  0.28 dB 0.54 dB  0.45 dB 1.7 dB	Synthesized Level Generator  Vista, CA Fremont, CA Orlando, FL



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Absolute Power Source Into 50 Ω SWR ≤ 1.6:1	(+10 to -10) dBm 10 MHz to 2 GHz	1.5 dB	Synthesized Level Generator  Vista, CA Fremont, CA Orlando, FL
	(2 to 20) GHz	1.7 dB	
SWR ≤ 1.6:1	(+5 to -10) dBm 10MHz to 2 GHz	0.82 dB	
SWR ≤ 1.8:1	(2 to 20) GHz	0.96 dB	
	(20 to 40) GHz	1.2 dB	
SWR ≤ 2.0:1	(+2.5 to -10) dBm (40 to 50) GHz	2.3 dB	
SWR ≤ 1.6:1	(-10 to -60) dBm 10MHz to 2 GHz	1.2 dB	
SWR ≤ 1.8:1	(2 to 20) GHz	1.3dB	
SWR ≤ 2.0:1	(20 to 40) GHz	1.6 dB	
	(40 to 50) GHz	2.3 dB	
SWR ≤ 1.6:1	(-60 to -110) dBm 10 MHz to 2 GHz	1.8 dB	
SWR ≤ 1.8:1	(2 to 20) GHz	1.9 dB	
SWR ≤ 2.0:1	(20 to 40) GHz	2.2 dB	
	(40 to 50) GHz	3.3 dB	
RF Absolute Power Source Into 50 Ω	(+24 to -48) dBm Up to 300 MHz	0.06 dB	Synthesized Low Phase Noise Level Generator  Vista, CA
	(+14 to -48) dBm 300 MHz to 3GHz	0.08 dB	
	(3 to 4) GHz	0.52 dB	
	(-48 to -84) dBm 100 kHz to 10 MHz	0.52 dB	
	(10 to 300) MHz	0.32 dB	
	300 MHz to 1.4 GHz	0.53 dB	
	(-48 to -74) dBm (1.4 to 4) GHz	0.53 dB	
	(-74 to -84) dBm (1.4 to 4) GHz	1 dB	
	(-84 to -94) dBm 100 kHz to 300 MHz	0.53 dB	
	300 MHz to 4 GHz	1 dB	
	(-94 to -130) dBm (10 to 128) MHz	0.73 dB	
	300 MHz to 4 GHz	1.6 dB	

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RF Absolute Power Source Into 75 Ω	(+14 to -23) dBm Up to 125 MHz (125 to 300) MHz 300 MHz to 1.4 GHz (1.4 to 3) GHz (3 to 4) GHz (-23 dBm to -54) dBm Up to 300 MHz 300 MHz to 4 GHz (-54 to -80) dBm 100 kHz to 300 MHz 300 MHz to 4 GHz (-80 to -90) dBm 100 kHz to 300 MHz 300 MHz to 4 GHz (-90 to -100) dBm 100 kHz to 300 MHz 300 MHz to 4 GHz (-100 to 120) dBm 10 MHz to 4GHz	0.07 dB 0.16 dB 0.26 dB 0.33 dB 0.52 dB 0.16 dB 0.52 dB 0.23 dB 0.54 dB 0.7 dB 1 dB 0.76 dB 1 dB 1.6 dB	Synthesized Low Phase Noise Level Generator  Vista, CA
RF Tuned Power – Measure (relative)	100 kHz to 4.2 GHz (+30 to 0) dB (0 to -30) dB (-30 to -60) dB (-60 to -90) dB (-90 to -129) dB (4.2 to 18) GHz (+30 to 0) dB (0 to -30) dB (-30 to -60) dB (-60 to -90) dB (-90 to -129) dB (18 to 26.5) GHz (+30 to 0) dB (0 to -30) dB (-30 to -60) dB (-60 to -90) dB (-90 to -129) dB	0.14 dB 0.16 dB 0.19 dB 0.21 dB 0.25 dB 0.18 dB 0.2 dB 0.22 dB 0.24 dB 0.28 dB 0.24 dB 0.25 dB 0.27 dB 0.29 dB 0.32 dB	Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
RF Tuned Power – Measure (relative)	(26.5 to 50) GHz (+30 to 0) dB (0 to -30) dB (-30 to -60) dB (-60 to -90) dB (-90 to -129) dB	0.2 dB 0.21 dB 0.23 dB 0.25 dB 0.29 dB	Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Digital Modulation - Measure Carrier Frequency 2 MHz to 2.65 GHz  Error Vector Magnitude for Modulation Types: MSK, GMSK, BPSK, DQPSK, Π/4DQPSK, 8PSK, 16QAM, 32 QAM and QPSK  Phase Error for Modulation Types: MSK, GMSK, BPSK, DQPSK, Π/4DQPSK, 8PSK, 16QAM, 32 QAM and QPSK	(0 to 15) % (1 to 100) kHz 100 kHz to 1 MHz 1 MHz to 2.65 GHz  (0 to 3)° (1 to 100) kHz (0.1 to 1) MHz 1 MHz to 2.65 GHz	0.33 % of reading 0.51 % of reading 1 % of reading  0.18° 0.34° 0.57°	Vector Signal Analyzer  Vista, CA Fremont, CA Orlando, FL
Digital Modulation - Measure Error Vector Magnitude for FSK Modulation	Modulation Frequency 3.2 kHz 1.152 kHz	0.54 % of reading 1.5 % of reading	Vector Signal Analyzer  Vista, CA Fremont, CA Orlando, FL
Amplitude Modulation - Source (11 to 13.5) MHz  20 Hz to 100 kHz  DC to 15 kHz	Rate:50 Hz to 50 kHz, (5 to 99) % Depth  Rate: (20 to 50) Hz (5 to 99) % Depth  Rate:9 kHz to 3.2 GHz (0 to 100) % Depth	0.1 % of reading  0.25 % of reading  5 % of reading	AM/FM Test Source, Analog Modulation Sources  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Amplitude Modulation - Source DC to 100 kHz	Rate: 250 kHz to 50 GHz (0 to 99) % Depth	1.2 % of reading + 0.07 AM	AM/FM Test Source, Analog Modulation Sources  Vista, CA Fremont, CA
Amplitude Modulation - Measure 100 kHz to 10 MHz  10 MHz to 3 GHz	Rate: 50 Hz to 10 kHz, (5 to 99) % Depth  50 Hz to 100 kHz, (20 to 99) % Depth	0.001 4 % of reading + 0.009 AM  0.001 % of reading + 0.007 AM	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Amplitude Modulation - Measure 10 MHz to 3 GHz  (3 to 26.5) GHz  (3 to 26.5) GHz	50 Hz to 100 kHz, (5 to 20) % Depth  50 Hz to 100 kHz, (5 to 20) % Depth  50 Hz to 100 kHz, (20 to 99) % Depth	0.001 % of reading + 0.029 AM  0.01 % of reading + 0.052 AM  0.01 % of reading + 0.018 AM	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Amplitude Modulation - Measure (26.5 to 31.15) GHz  (31.15 to 50) GHz	50 Hz to 100kHz, (5 to 20) % Depth (20 to 99) % Depth  50 Hz to 100kHz, (5 to 20) % Depth (20 to 99) % Depth	0.08 % of reading + 0.000 2 AM 0.02 % of reading + 0.000 14 AM  0.3 % of reading + 0.000 05 AM 0.07 % of reading + 0.000 046 AM	Microwave Measuring Receiver System  Vista, CA

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Amplitude Modulation - Measure 150 kHz to 10 MHz Rate: 50 Hz to 10 kHz, Depth: 5 % to 99 % Rate: 20 Hz to 10 kHz, Depth: to 99 % 10 MHz to 1.3 GHz Rate: 50 Hz to 50 kHz, Depth: 5 % to 99 % Rate: 20 Hz to 10 kHz, Depth: to 99 % (1.3 to 26.5) GHz Rate: 50 Hz to 10 kHz, Depth: 5 % to 99 % 10 MHz to 26.5 GHz Rate: 20 Hz to 10 kHz, Depth: to 99 %	(5 to < 10) % Depth (10 to 99) % Depth (5 to < 10) % Depth (10 to 99) % Depth  (5 to < 10) % Depth (10 to 99) % Depth (5 to < 10) % Depth (10 to 99) % Depth  (5 to < 10) % Depth (10 to 99) % Depth  (5 to < 10) % Depth (10 to 99) % Depth	0.023 AM + 0.01 % of reading 0.023 AM + 0.06 % of reading 0.035 AM + 0.01 % of reading 0.035 AM + 0.06 % of reading  0.012 AM + 0.01 % of reading 0.012 AM + 0.06 % of reading 0.035 AM + 0.01 % of reading 0.035 AM + 0.06 % of reading  0.018 AM + 0.03 % of reading 0.018 AM + 0.07 % of reading  0.035 AM + 0.03 % of reading 0.035 AM + 0.07 % of reading	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Measure 250 kHz to 10 MHz  10 MHz to 6.6 GHz	Rate: 20 Hz to 10 kHz Dev:20 Hz to 40 kHz pk  Rate: 50 Hz to 200 kHz Dev:250 Hz to 400 kHz pk	1.7 % of reading + 5.7 Hz 1.1 % of reading + 6.6 Hz  1.8 % of reading + 5.1 Hz 1.2 % of reading + 6.1 Hz	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Measure (6.6 to 13.2) GHz  (13.2 to 26.5) GHz	Rate: 50 Hz to 200 kHz Dev:250 Hz to 400 kHz pk  Rate: 50 Hz to 100 kHz Dev:250 Hz to 400 kHz pk	2.9 % of reading + 4 Hz 1.2 % of reading + 6.4 Hz  4.4 % of reading + 3.8Hz 1.2 % of reading + 7.6 Hz	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Measure (26.5 to 50.0) GHz	Rate: 20 Hz to 10 kHz Dev:20 Hz to 40 kHz pk	10 % of reading + 4.2 Hz 1 % of reading + 12 Hz	Microwave Measuring Receiver System  Vista, CA
Frequency Modulation - Measure 250 kHz to 10 MHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM 10 MHz to 26.5 GHz	Rate: 20 Hz to 10 kHz ≤ 40 kHz pk	0.024 FM + 2.4 Hz Pk 0.024 FM + 8 Hz Pk	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation - Measure 10 MHz to 1.3 GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM	Rate: 50 Hz to 100 kHz ≤ 400 kHz pk	0.012 FM + 2.4 Hz Pk 0.012 FM + 14 Hz Pk 0.012 FM + 66 Hz Pk	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Measure (>1.3 to 6.2) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM (> 6.2 to 12.4) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM	Rate: 50 Hz to 100 kHz ≤ 400 kHz pk	0.012 FM + 19 Hz Pk 0.012 FM + 23 Hz Pk 0.012 FM + 75 Hz Pk  0.012 FM + 35 Hz Pk 0.012 FM + 39 Hz Pk 0.012 FM + 91 Hz Pk	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Measure (> 12.4 to 18.6) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM (> 18.6 to 26.5) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM	Rate: 50 Hz to 100 kHz ≤ 400 kHz pk	0.012 FM + 51 Hz Pk 0.012 FM + 55 Hz Pk 0.012 FM + 110 Hz Pk  0.012 FM + 67 Hz Pk 0.012 FM + 71 Hz Pk 0.012 FM + 130 Hz Pk	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Measure 10 MHz to 26.5 GHz 10 MHz to 1.3 GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM (>1.3 to 6.2) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM	Rate: 50 Hz to 100 kHz ≤ 400 kHz pk	0.058 FM + 3.9 Hz Pk 0.058 FM + 8 Hz Pk 0.058 FM + 66 Hz Pk  0.058 FM + 19 Hz Pk 0.058 FM + 23 Hz Pk 0.058 FM + 75 Hz Pk	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency Modulation - Measure (> 6.2 to 12.4) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM (> 12.4 to 18.6) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM (> 18.6 to 26.5) GHz (0 to < 4) kHz pk FM (≥ 4 to < 40) kHz pk FM (≥ 40 to < 400) kHz pk FM	Rate: 20 Hz to 200 kHz ≤ 400 kHz pk	0.058 FM + 35 Hz Pk 0.058 FM + 39 Hz Pk 0.058 FM + 91 Hz Pk  0.058 FM + 51 Hz Pk 0.058 FM + 55 Hz Pk 0.058 FM + 110 Hz Pk  0.058 FM + 67 Hz Pk 0.058 FM + 71 Hz Pk 0.058 FM + 130 Hz Pk	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Source (11 to 13.5) MHz  (11 to 13.5) MHz  (88 to 108) MHz  (88 to 108) MHz  (352 to 432) MHz	Rate: DC to 10 kHz Dev.: ≤ 100 kHz peak  Rate: DC to 10 kHz Dev.: ≤ 200 kHz peak  Rate: DC to 100 kHz Dev.: ≤ 100 kHz peak  Rate: DC to 100 kHz Dev.: ≤ 200 kHz peak  Rate: DC to 100 kHz Dev.: ≤ 100 kHz peak	0.1 % of reading  0.25 % of reading  0.1 % of reading  0.25 % of reading  0.1 % of reading	AM/FM Test Source  Vista, CA Fremont, CA Orlando, FL
Frequency Modulation - Source (352 to 432) MHz  9 kHz to 1 GHz  (1 to 2) GHz  (2 to 3) GHz	Rate: DC to 100 kHz Dev.: ≤ 200 kHz peak  Rate: DC to 150 kHz Dev.: ≤ 200 kHz peak  Rate: DC to 150 kHz Dev.: ≤ 400 kHz peak  Rate: DC to 150 kHz Dev.: ≤ 400 kHz peak	0.25 % of reading  3 % of reading + 30 Hz  3 % of reading + 60 Hz  3 % of reading + 120 Hz	Analog Modulation Sources  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Frequency Modulation - Source 1 kHz rate Max. Dev. 2 MHz Max. Dev. 4 MHz Max. Dev. 8 MHz Max. Dev. 16 MHz Max. Dev. 32 MHz Max. Dev. 64 MHz Max. Dev. 128 MHz	250 kHz to 1 GHz (1 to 2) GHz (2 to 3.2) GHz (3.2 to 10) GHz (10 to 20) GHz (20 to 40) GHz (40 to 50) GHz	40 mHz/Hz + 23 Hz	Analog Modulation Sources  Vista, CA Fremont, CA Orlando, FL
Phase Modulation - Measure >0.7 rad Dev. >0.6 rad Dev. >1.2 rad Dev.	100 kHz to 6.6 GHz (6.6 to 13.2) GHz (13.2 to 26.5) GHz	1.2 % of reading + 0.0071 rad 1.2 % of reading + 0.0071 rad 1.2 % of reading + 0.0073 rad	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Phase Modulation - Measure >1.3 Rad Dev. >2.4 Rad Dev.	(26.5 to 31.5 GHz) (31.5 to 50) GHz	1 % of reading + 0.0073 rad 1 % of reading + 0.008 rad	Microwave Measuring Receiver System  Vista, CA
Phase Modulation - Measure 150 kHz to 10 MHz (0 to < 4) rad pk PM (≥ 4 to < 40) rad pk PM 10 MHz to 1.3 GHz (0 to < 4) rad pk PM (≥ 4 to < 40) rad pk PM (≥ 40 to < 400) rad pk PM	200 Hz to 10 kHz Rate ≤ 40 rad pk  200 Hz to 20 kHz Rate ≤ 400 rad pk	4.7 % of reading + 0.001 rad 4.7 % of reading + 0.008 rad  3.5 % of reading + 0.001 rad 3.5 % of reading + 0.008 rad 3.5 % of reading + 0.09 rad	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Phase Modulation - Measure 1.3 to 26.5 GHz (0 to < 4) rad pk PM (≥ 4 to < 40) rad pk PM (≥ 40 to < 400) rad pk PM	200 Hz to 20 kHz Rate ≤ 400 rad pk	3.5 % of reading + 0.001 rad 3.5 % of reading + 0.008 rad 3.5 % of reading + 0.09 rad	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL



ANSI National Accreditation Board

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Phase Modulation - Source Rate: 20 Hz to 10 kHz Carrier: 9 kHz to 1 GHz (1 to 2) GHz (2 to 3.2) GHz  250 kHz to 1 GHz (>1 to 2) GHz (>2 to 3.2) GHz (>3.2 to 10) GHz (>10 to 20) GHz (>20 to 40) GHz (>40 to 50) GHz	(0 to 10) rad (0 to 20) rad (0 to 40) rad  Max. Dev. (0 to 20) rad (0 to 40) rad (0 to 80) rad (0 to 160) rad (0 to 320) rad (0 to 640) rad (0 to 1 280) rad	0.05 rad + 3 % of reading 0.1 rad + 3 % of reading 0.2 rad + 3 % of reading  0.012 rad + 6 % of reading 0.012 rad + 6 % of reading	Analog Modulation Sources  Vista, CA Fremont, CA Orlando, FL
Distortion – Measure	(0.01 to 100) % Distortion 250 kHz to 50 GHz	0.065 % of reading	Microwave Measuring Receiver System  Vista, CA Fremont, CA Orlando, FL
Distortion - Measure Fundamental Frequency 20 Hz to 20 kHz (20 to 100) kHz	(-99 to 0) dB (-99 to 0) dB	1.2 dB 2.3 dB	Distortion Analyzer  Vista, CA Fremont, CA Orlando, FL
Power Reference- Measure 50 MHz	1 mW	5.5 $\mu$ W	Primary Standards Lab (H75) Thermistor Mount, Power Meter, Precision 8.5-digit Multimeter  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Absolute Power - Measure	9 kHz to 4.2 GHz (+20 to -30) dBm	0.3 % of Reading + 0.65 dB	RF Power Meters/Sensors  Vista, CA Fremont, CA Orlando, FL
	10 MHz to 18 GHz (-20 to -70) dBm	0.44 % of reading + 0.32 dB	
	50 MHz to 26.5 GHz (+20 to -30) dBm	0.24 % of reading + 0.86 dB	
	(26.5 to 50) GHz (+20 to -30) dBm	0.46 % of reading + 0.94 dB	
	(-20 to -70) dBm	1.6 % of reading + 0.67 dB	
Power Meter – Range Calibration	3 $\mu$ W	21 nW	Power Meter Range Calibrator w/ Precision DC Voltage Source  Vista, CA Fremont, CA Orlando, FL
	10 $\mu$ W	19 nW	
	30 $\mu$ W	21 nW	
	100 $\mu$ W	77 nW	
	300 $\mu$ W	0.11 $\mu$ W	
	1 mW	0.29 $\mu$ W	
	3 mW	0.64 $\mu$ W	
	10 mW	6.4 $\mu$ W	
	30 mW	13 $\mu$ W	
	100 mW	0.10 mW	
Noise Figure - Source	15 dB ENR		Primary Standards Lab Noise Source  Vista, CA
	10 MHz to 1.5 GHz	0.29 dB	
	(1.5 to 3) GHz	0.25 dB	
	(3 to 7) GHz	0.26 dB	
	(7 to 18) GHz	0.37 dB	
Noise Figure - Source	(18 to 26.5) GHz	0.42 dB	Noise Source  Fremont, CA Orlando, FL
	15 dB ENR		
	10 MHz to 4 GHz	0.22 dB	
	(>4 to 10) GHz		
Noise Figure - Measure	(>10 to 18) GHz		Noise Figure Measurement System w/ Standard Noise Sources  Vista, CA Fremont, CA Orlando, FL
	(>18 to 26.5) GHz		
	100 kHz to 30 MHz	0.43 dB	
	>30 MHz to 3 GHz	0.42 dB	
	(>3 to 26.5) GHz	0.47 dB	

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Single Sideband Phase Noise - Measure	1 MHz to 26.5 GHz	1.8 dB	Phase Noise Measurement System  Vista, CA Fremont, CA Orlando, FL
Single Sideband Phase Noise - Measure	(26.5 to 50) GHz	1.8 dB	Phase Noise Measurement System  Vista, CA
Power Sensor Calibration Factor	(-60 to < 0 > to 23) dBm 9 kHz to 10 MHz >10 MHz to 4 GHz (>4 to 10 GHz) >10 to 18 GHz >18 to 26.5 GHz >26.5 to 44 GHz >44 to 50 GHz	1.3 % of reading 1.4 % of reading 1.7 % of reading 1.9 % of reading 3.8 % of reading 5.1 % of reading 7.6 % of reading	Signal Sources, Precision Level Source and Tegam System  Vista, CA
Current Probes and Bulk Current Injection Probes Insertion Loss	(0 to 1) dB 20 Hz to 300 kHz 300 kHz to 400 MHz	0.09 dB 0.09 dB	Network/RF Impedance Analyzers  Vista, CA Fremont, CA Orlando, FL
Transfer Impedance	(0 to 1) dB 20 Hz to 300 kHz 300 kHz to 400 MHz	0.02 dB 0.02 dB	



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Transmission $S_{12}/S_{21}$ - Measure Phase and Magnitude	9 kHz to 1.3 GHz (-180 to 180) <sup>o</sup>		LF Vector Network Analyzer, Calibration Kits  Vista, CA Fremont, CA Orlando, FL
	(0 to 10) dB	0.18 dB (1 <sup>o</sup> )	
	(-10 to 0) dB	0.057 dB (0.43 <sup>o</sup> )	
	(-20 to -10) dB	0.072 dB (0.8 <sup>o</sup> )	
	(-60 to -20) dB	0.086 dB (0.89 <sup>o</sup> )	
	(-70 to -60) dB	0.13 dB (1.2 <sup>o</sup> )	
	(-80 to -70) dB	0.3 dB (2.4 <sup>o</sup> )	
	(-90 to -80) dB	0.92 dB (6.9 <sup>o</sup> )	
	(1.3 to 3) GHz		
	(0 to 10) dB	0.12 dB (4.1 <sup>o</sup> )	
	(-10 to 0) dB	0.063 dB (0.48 <sup>o</sup> )	
	(-20 to -10) dB	0.077 dB (0.8 <sup>o</sup> )	
	(-60 to -20) dB	0.093 dB (0.94 <sup>o</sup> )	
	(-70 to -60) dB	0.13 dB (1.2 <sup>o</sup> )	
Transmission $S_{12}/S_{21}$ - Measure Phase and Magnitude	(1.3 to 3) GHz		Vector Network Analyzer w/ Calibration Kits  Vista, CA Fremont, CA Orlando, FL
	(-80 to -70) dB	0.32 dB (2.6 <sup>o</sup> )	
	(-90 to -80) dB	0.92 dB (7.4 <sup>o</sup> )	
	50 MHz to 2 GHz (-180 to 180) <sup>o</sup>		
	(0 to 10) dB	0.07 dB (0.46 <sup>o</sup> )	
	(-10 to 0) dB	0.054 dB (0.36 <sup>o</sup> )	
	(-20 to -10) dB	0.075 dB (0.5 <sup>o</sup> )	
	(-30 to -20) dB	0.12 dB (0.83 <sup>o</sup> )	
	(-40 to -30) dB	0.29 dB (1.9 <sup>o</sup> )	
	(-50 to -40) dB	2.3 dB (5.5 <sup>o</sup> )	
	(-60 to -50) dB	5.9 dB (5.7 <sup>o</sup> )	
	(2 to 8) GHz		
	(0 to 10) dB	0.09 dB (0.59 <sup>o</sup> )	
	(-10 to 0) dB	0.07 dB (0.46 <sup>o</sup> )	
(-20 to -10) dB	0.087 dB (0.57 <sup>o</sup> )		
(-30 to -20) dB	0.1 dB (0.69 <sup>o</sup> )		
(-40 to -30) dB	0.12 dB (0.81 <sup>o</sup> )		
(-50 to -40) dB	0.15 dB (0.98 <sup>o</sup> )		
(-60 to -50) dB	0.22 dB (1.4 <sup>o</sup> )		
(-70 to -60) dB	0.45 dB (3.1 <sup>o</sup> )		
(-80 to -70) dB	1.2 dB (8.7 <sup>o</sup> )		
(-90 to -80) dB	3.4 dB (8.7 <sup>o</sup> )		

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Transmission $S_{12}/S_{21}$ - Measure Phase and Magnitude	(8 to 20) GHz (-180 to 180° (0 to 10) dB (-10 to 0) dB (-20 to -10) dB (-30 to -20) dB (-40 to -30) dB (-50 to -40) dB (-60 to -50) dB (-70 to -60) dB (-80 to -70) dB (-90 to -80) dB	0.19 dB (1.3°) 0.17 dB (1.1°) 0.18 dB (1.2°) 0.2 dB (1.3°) 0.22 dB (1.4°) 0.24 dB (1.6°) 0.32 dB (2.1°) 0.6 dB (4.1°) 1.6 dB (8.7°) 4.2 dB (11°)	Vector Network Analyzer w/ Calibration Kits  Vista, CA Fremont, CA Orlando, FL
	(20 to 50) GHz (0 to 10) dB (-10 to 0) dB (-20 to -10) dB (-30 to -20) dB (-40 to -30) dB (-50 to -40) dB (-60 to -50) dB (-70 to -60) dB (-80 to -70) dB (-90 to -80) dB	0.54 dB (3.7°) 0.5 dB (3.4°) 0.49 dB (3.3°) 0.5 dB (3.4°) 0.52 dB (3.5°) 0.55 dB (3.7°) 0.65 dB (4.4°) 1.1 dB (7.6°) 2.7 dB (8.9°) 6.6 dB (11°)	
Transmission $S_{12}/S_{21}$ – Measure  Magnitude  Phase	(-90 to 10) dB (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz $\leq 0$ to $\leq + 60^\circ$ 0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz	0.014 dB 0.043 dB 0.11 dB 0.14 dB  0.023° 0.013° 0.064° 0.017°	Vector Network Analyzer w/ Calibration Kits  Vista, CA

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
<p style="text-align: center;">Reflection <math>S_{11}/S_{22}</math> - Measure Phase and Magnitude <sup>2,3</sup></p>	<p>9 kHz to 1.3 GHz (-180 to 180)<sup>o</sup></p> <p style="padding-left: 40px;">(&lt;0.03) <math>\Gamma</math> (&lt;0.2) <math>\Gamma</math> (&lt;0.4) <math>\Gamma</math> (&lt;0.6) <math>\Gamma</math> (&lt;0.8) <math>\Gamma</math> (&lt;1) <math>\Gamma</math></p> <p>(1.3 to 3) GHz</p> <p style="padding-left: 40px;">(&lt;0.03) <math>\Gamma</math> (&lt;0.2) <math>\Gamma</math> (&lt;0.4) <math>\Gamma</math> (&lt;0.6) <math>\Gamma</math> (&lt;0.8) <math>\Gamma</math> (&lt;1) <math>\Gamma</math></p>	<p style="text-align: center;">0.002 3 (1.4<sup>o</sup>) 0.003 5 (1.4<sup>o</sup>) 0.006 1 (1.1<sup>o</sup>) 0.008 3 (0.94<sup>o</sup>) 0.01 (0.83<sup>o</sup>) 0.012 (0.7<sup>o</sup>)</p> <p style="text-align: center;">0.005 6 (1.8<sup>o</sup>) 0.003 5 (1.8<sup>o</sup>) 0.008 (1.3<sup>o</sup>) 0.011 (1.2<sup>o</sup>) 0.013 (1<sup>o</sup>) 0.016 (0.89<sup>o</sup>)</p>	<p style="text-align: center;">LF Vector Network Analyzer, Calibration Kits</p> <p style="text-align: center;">Vista, CA Fremont, CA Orlando, FL</p>
<p style="text-align: center;">Reflection <math>S_{11}/S_{22}</math> – Measure Phase and Magnitude <sup>2,3</sup></p>	<p>50 MHz to 2 GHz (-180 to 180)<sup>o</sup></p> <p style="padding-left: 40px;">(&lt;0.03) <math>\Gamma</math> (&lt;0.2) <math>\Gamma</math> (&lt;0.4) <math>\Gamma</math> (&lt;0.6) <math>\Gamma</math> (&lt;0.8) <math>\Gamma</math> (&lt;1) <math>\Gamma</math></p> <p>(2 to 8) GHz</p> <p style="padding-left: 40px;">(&lt;0.03) <math>\Gamma</math> (&lt;0.2) <math>\Gamma</math> (&lt;0.4) <math>\Gamma</math> (&lt;0.6) <math>\Gamma</math> (&lt;0.8) <math>\Gamma</math> (&lt;1) <math>\Gamma</math></p> <p>(8 to 20) GHz</p> <p style="padding-left: 40px;">(&lt;0.03) <math>\Gamma</math> (&lt;0.2) <math>\Gamma</math> (&lt;0.4) <math>\Gamma</math> (&lt;0.6) <math>\Gamma</math> (&lt;0.8) <math>\Gamma</math> (&lt;1) <math>\Gamma</math></p>	<p style="text-align: center;">0.01 (3.3<sup>o</sup>) 0.012 (3.3<sup>o</sup>) 0.014 (2<sup>o</sup>) 0.017 (1.6<sup>o</sup>) 0.021 (1.5<sup>o</sup>) 0.026 (1.5<sup>o</sup>)</p> <p style="text-align: center;">0.01 (3.5<sup>o</sup>) 0.012 (3.5<sup>o</sup>) 0.015 (2.2<sup>o</sup>) 0.019 (1.8<sup>o</sup>) 0.024 (1.7<sup>o</sup>) 0.003 (1.7<sup>o</sup>)</p> <p style="text-align: center;">0.002 4 (7.8<sup>o</sup>) 0.002 7 (7.8<sup>o</sup>) 0.003 3 (4.8<sup>o</sup>) 0.004 2 (4<sup>o</sup>) 0.053 (3.8<sup>o</sup>) 0.067 (3.9<sup>o</sup>)</p>	<p style="text-align: center;">Vector Network Analyzer w/ Calibration Kits</p> <p style="text-align: center;">Vista, CA Fremont, CA Orlando, FL</p>

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Reflection $S_{11}/S_{22}$ – Measure Phase and Magnitude <sup>2,3</sup>	(20 to 50) GHz (-180 to 180)° (<0.03) $\Gamma$ (<0.2) $\Gamma$ (<0.4) $\Gamma$ (<0.6) $\Gamma$ (<0.8) $\Gamma$ (<1) $\Gamma$	0.06 (19°) 0.066 (11°) 0.079 (11°) 0.097 (9.3°) 0.12 (8.8°) 0.15 (8.8°)	Vector Network Analyzer w/ Calibration Kits  Vista, CA Fremont, CA Orlando, FL
Reflection $S_{11}/S_{22}$ – Measure Magnitude <sup>2,3</sup>  Phase	(<0.03 to <1) $\Gamma$ (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz </= 0 to </= 60° (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz	0.011 $\Gamma$ 0.011 $\Gamma$ 0.018 $\Gamma$ 0.023 $\Gamma$  0.023° 0.013° 0.064° 0.017°	Vector Network Analyzer w/ Calibration Kits  Vista, CA
Network Analyzer System Verification (Corrected Performance) 7mm Test Ports Reflection $S_{11}/S_{22}$ – Magnitude <sup>2,3</sup>  Phase	</= 0.1 to </= 1.0 $\Gamma$ (0.0003 to 1) GHz (1 to 3) GHz (3 to 6) GHz </= 0 to </= 60° (0.0003 to 1) GHz (1 to 3) GHz (3 to 6) GHz	0.003 8 $\Gamma$ 0.003 9 $\Gamma$ 0.005 1 $\Gamma$  2° 2.1° 2.7°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL
Network Analyzer System Verification (Corrected Performance) N-Type Test Ports Reflection $S_{11}/S_{22}$ – Magnitude <sup>2,3</sup>  Phase	</= 0.1 to </= 1.0 $\Gamma$ (0.0003 to 1) GHz (1 to 3) GHz (3 to 8) GHz (8 to 18) GHz </= 0 to </= 60° (0.000 3 to 1) GHz (1 to 3) GHz (3 to 8) GHz (8 to 18) GHz	0.009 $\Gamma$ 0.001 $\Gamma$ 0.001 5 $\Gamma$ 0.004 1 $\Gamma$  0.46° 0.64° 0.65° 1.3°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL



ANSI National Accreditation Board

Electrical – RF/Microwave

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Network Analyzer System Verification (Corrected Performance) 3.5mm Test Ports Reflection $S_{11}/S_{22}$ – Magnitude <sup>2,3</sup>  Phase	$\leq 0.1$ to $\leq 1.0 \Gamma$ (0.045 to 2) G $\Gamma$ Hz (2 to 20) GHz (20 to 26.5) GHz  $\leq 0$ to $\leq 60^\circ$ (0.045 to 2) GHz (2 to 20) GHz (20 to 26.5) GHz	0.000 76 $\Gamma$ 0.001 $\Gamma$ 0.003 $\Gamma$  0.34° 0.36° 0.64°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL
Network Analyzer System Verification (Corrected Performance) 2.4mm Test Ports Reflection $S_{11}/S_{22}$ – Magnitude <sup>2,3</sup>  Phase	$\leq 0.1$ to $\leq 1.0 \Gamma$ (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz  $\leq 0$ to $\leq 60^\circ$ (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz	0.003 8 $\Gamma$ 0.003 8 $\Gamma$ 0.004 4 $\Gamma$ 0.005 5 $\Gamma$  0.96° 1.3° 1.5° 2.3°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL
Network Analyzer System Verification (Corrected Performance) 7mm Test Ports Transmission $S_{21}/S_{12}$ – Magnitude  Phase	$\leq 0.1$ to $\leq 1.0$ dB (0.0003 to 1) GHz (1 to 3) GHz (3 to 6) GHz  $\leq 0$ to $\leq 60^\circ$ (0.0003 to 1) GHz (1 to 3) GHz (3 to 6) GHz	0.16 dB 0.16 dB 0.17 dB  1.4° 1.5° 2.2°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Network Analyzer System Verification (Corrected Performance) N-Type Test Ports Transmission $S_{21}/S_{12}$ – Magnitude  Phase	$\leq 0.1$ to $\leq 1.0$ dB (0.0003 to 1) GHz (1 to 3) GHz (3 to 8) GHz (8 to 18) GHz  $\leq 0$ to $\leq 60^\circ$ (0.0003 to 1) GHz (1 to 3) GHz (3 to 8) GHz (8 to 18) GHz	0.018 dB 0.019 dB 0.02 dB 0.024 dB  0.18° 0.2° 0.23° 0.24°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL
Network Analyzer System Verification (Corrected Performance) 3.5mm Test Ports Transmission $S_{21}/S_{12}$ – Magnitude  Phase	$\leq 0.1$ to $\leq 1.0$ dB (0.045 to 2) GHz (2 to 20) GHz (20 to 26.5) GHz  $\leq 0$ to $\leq 60^\circ$ (0.045 to 2) GHz (2 to 20) GHz (20 to 26.5) GHz	0.018 dB 0.019 dB 0.032 dB  0.18° 0.18° 0.32°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL
Network Analyzer System Verification (Corrected Performance) 2.4mm Test Ports Transmission $S_{21}/S_{12}$ – Magnitude  Phase	$\leq 0.1$ to $\leq 1.0$ dB (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz  $\leq 0$ to $\leq 60^\circ$ (0.045 to 2) GHz (2 to 20) GHz (20 to 40) GHz (40 to 50) GHz	0.023 dB 0.033dB 0.055 dB 0.068 dB  0.23° 0.41° 0.97° 1.8°	Mechanical Calibration Kits and RF Network Analyzer System  Vista, CA Fremont, CA Orlando, FL
Scalar Network Analyzer  Absolute Log Error  Dynamic Accuracy	21 dB  (0 to 140) dB	0.04 dB  0.02 dB	Scalar Network Analyzer Calibrator  Vista, CA Fremont, CA Orlando, FL

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Network Analyzer Dynamic Accuracy	(0 to 140) dB	0.02 dB	Dynamic Accuracy Test Set Vista, CA Fremont, CA Orlando, FL
Impulse Spectral Amplitude Source CISPR Band A CISPR Band B CISPR Band C and D CISPR Band E	(10 to 150) kHz 150 kHz to 30 MHz 30 MHz to 1 GHz (1 to 18) GHz	0.82 dB 0.82 dB 1.1 dB 1.5 dB	Pulse Generator  Vista, CA Fremont, CA Orlando, FL
Sinewave Output for CISPR Checks Source @ 60 dB/ $\mu$ V	(0 to -70) dB 100 kHz (1, 10, and 100) MHz	0.35 dB 0.35 dB	
Peak and Average Detector Response	(0 to -70) dB CISPR Band A thru D	1.3 dB	
LISN (Line Impedance Stabilization Network) & AMN (Artificial Mains Network) Insertion Loss	(-20 to 0) dB 10 Hz to 1000 MHz	0.25 dB	RF Impedance/Network Analyzers, Calibration Kits
Isolation (De-Coupling Factor)	(-90 to 0) dB 10 Hz to 1 000 MHz	2.3 dB	Vista, CA Fremont, CA Orlando, FL
Impedance (Magnitude)	0.1 $\Omega$ to 1 k $\Omega$ 10 Hz to 1000 MHz	2.4 % of reading	
Impedance (Phase)	(-180 to 180) $^{\circ}$ 10 Hz to 1 000 MHz	2.6 $^{\circ}$	

**Electrical – RF/Microwave**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
CDN (Coupling-Decoupling Network) & ISN (Impedance Stabilization Network) Insertion Loss	(-20 to 0) dB 10 Hz to 1000 MHz	0.25 dB	RF Impedance/Network Analyzers, Calibration Kits  Vista, CA Fremont, CA Orlando, FL
Isolation (De-Coupling Factor)	(-90 to 0) dB 10 Hz to 1000 MHz	2.3 dB	
Impedance (Magnitude)	0.1 Ω to 1 kΩ 10 Hz to 1 000 MHz	2.4 % of reading	
Impedance (Phase)	(-180 to 180)° 10 Hz to 1 000 MHz	2.6°	
LCL (Longitudinal Conversion Loss)	(-20 to 0) dB 10 Hz to 1 000 MHz	2.3 dB	

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Test, Dial Indicators & Thickness Gages <sup>1</sup>	Up to 4 in	0.26 μin/in + 57 μin	Gage Blocks  Vista, CA Fremont, CA Orlando, FL
Calipers <sup>1</sup>	Up to 54 in	46 μin/in + 52 μin	Gage Blocks, Standard Rings, Rod Sets  Vista, CA Fremont, CA Orlando, FL
Micrometers <sup>1</sup> (Linearity Only)	Up to 54 in	46 μin/in + 52 μin	Gage Blocks, Rod Sets  Vista, CA Fremont, CA Orlando, FL

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pin and Plug Gages <sup>1</sup>	Up to 25 mm	5 μm	Micrometer Vista, CA Fremont, CA
Pin and Plug Gages	Up to 2.36 in	180 μin	Laser Micrometer & Pin Gages Vista, CA
Height Gages	Up to 12 in	(750 + 4.7L) μin	Gage Blocks Vista, CA Fremont, CA Orlando, FL
Depth Gages	Up to 12 in	(160 + 4.1L) μin	
Surface plates Overall Flatness Local Area Flatness	Up to 161 DL Up to 0.001 in	(77 + 0.18 DL) μin 29 μin	Leveling System Repeat-O-Meter Fremont, CA
Steel Rulers	Up to 48 in	0.036 in	Gage Blocks Vista, CA Fremont, CA Orlando, FL
Tape Measures	Up to 50 ft	0.036 in	Gage Blocks Vista, CA Fremont, CA Orlando, FL
Protractor/Angle	Up to 90°	0.069°	Angle Gage Blocks Vista, CA Fremont, CA Orlando, FL
Thread Plug Gage - Minor Diameter Major Diameter Pitch Diameter Flank Angle	(0.04 to 5.9) in (0.04 to 5.9) in (0.04 to 5.9) in (≥ 27 to ≤ 80)°	(76 + 5.8L) μin (76 + 5.8L) μin (40 + 7.7L) μin 0.11°	IAC MasterScanner Fremont, CA

**Length – Dimensional Metrology**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Ring Gage - Minor Diameter Major Diameter Pitch Diameter Flank Angle	(0.10 to 6.3) in (0.10 to 6.3) in (0.10 to 6.3) in ( $\geq 27$ to $\leq 80$ )°	(90 + 4.8L) $\mu$ in (90 + 4.8L) $\mu$ in (40 + 8.5L) $\mu$ in 0.11°	IAC MasterScanner  Fremont, CA

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Barometric Pressure Measurement	Up to 39 psia	0.014 psi	Digital Barometer/Precision Absolute Manometer  Vista, CA Fremont, CA Orlando, FL
Force-Measuring Equipment Tension / Compression	(0.1 to 50) kgf	0.042 mgf/gf + 7.1 gf	NIST Class F Weights  Vista, CA Fremont, CA Orlando, FL
Force-Measuring Equipment Tension / Compression	Up to 1 000 lbf (1 000 to 10 000) lbf (10 000 to 20 000) lbf (20 000 to 30 000) lbf (30 000 to 40 000) lbf (40 000 to 50 000) lbf	5.8 lbf 59 lbf 71 lbf 88 lbf 110 lbf 130 lbf	Reference Load Cells  Vista, CA Orlando, FL
Force-Measuring Standards/Transducers Tension / Compression	Up to 1 000 lbf (>1 000 to 5 000) lbf (>5 000 to 10 000) lbf (>10 000 to 25 000) lbf (>25 000 to 50 000) lbf (>50 000 to 75 000) lbf (>75 000 to 100 000) lbf	1.5 lbf 7.3 lbf 15 lbf 37 lbf 77 lbf 120 lbf 150 lbf	Reference Load Cells  Fremont, CA

**Mass and Mass Related**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Pressure – Source/Pressure gages and transducers <sup>1</sup>	(-12 to 300) psi (300 to 10 000) psi	0.2 psi 9.5 psi	Pressure Calibration System  Vista, CA Fremont, CA Orlando, FL
Torque – Measure/Torque tools <sup>1</sup>	(2.5 to 25) lbf·in (5 to 10) lbf·in (10 to 100) lbf·in (5 to 50) lbf·ft (25 to 250) lbf·ft (250 to 600) lbf·ft	0.03 lbf·in + 0.5 % of reading 0.008 7 lbf·in + 0.5 % of reading 0.067 lbf·in + 0.5 % of reading 0.002 9 lbf·ft + 0.6 % of reading 0.17 lbf·ft + 0.5 % of reading 0.15 lbf·ft + 0.6 % of reading	Torque Transducers  Vista, CA Fremont, CA
Torque – Measure/Torque tools	(5 to 50) lbf·in (50 to 500) lbf·in (10 to 100) lbf·ft (100 to 750) lbf·ft	0.016 lbf·in + 0.003 lbf·in/lbf·in 0.068 lbf·in + 0.002 lbf·in/lbf·in 0.39 lbf·ft + 0.001 lbf·ft/lbf·ft 2.9 lbf·ft + 0.001 lbf·ft/lbf·ft	Torque Transducers  Orlando, FL
Torque – Source/Analyzers and measuring equipment	(1 to 10) lbf·in (10 to 25) lbf·in (25 to 50) lbf·in (50 to 100) lbf·in (100 to 150) lbf·in (150 to 250) lbf·in	0.021 % of reading + 0.009 lbf·in 0.017 % of reading + 0.011 lbf·in 0.033 % of reading + 0.022 lbf·in 0.035 % of reading + 0.040 lbf·in 0.038 % of reading + 0.042 lbf·in 0.04 % of reading + 0.045 lbf·in	Calibration Wheels Standard Weights  Vista, CA Fremont, CA Orlando, FL
Scales & Balances <sup>1</sup>	1 mg to 10 g (10 to 500) g (0.5 to 1) kg (1 to 2) kg (2 to 20) kg (20 to 40) kg (40 to 50) kg	2.3 mg + 0.6R 81 mg + 0.6R 1.5 mg + 0.6R 2.9 mg + 0.6R 2.3 g + 0.6R 4.6 g + 0.6R 5.8 g + 0.6R	NIST Class F Weights  Vista, CA Fremont, CA
Scales & Balances <sup>1</sup>	(0.001 to 0.05) lb (0.1 to 1) lb (1 to 2) lb (2 to 5) lb (5 to 10) lb (10 to 20) lb (20 to 50) lb	0.000 012 lb + 0.6R 0.000 18 lb + 0.6R 0.000 24 lb + 0.6R 0.000 6 lb + 0.6R 0.001 1 lb + 0.6R 0.002 4 lb + 0.6R 0.006 lb + 0.6R	NIST Class F Weights  Vista, CA Fremont, CA

**Mass and Mass Related**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances <sup>4</sup>	0.25 oz 0.5 oz 1 oz 2 oz 4 oz 8 oz 16 oz 1 kg 2 kg 5 lb (1 to 10) lb (10 to 50) lb (50 to 300) lb	0.000 71 oz 0.001 1 oz 0.000 27 oz 0.000 46 oz 0.001 oz 0.002 2 oz 0.002 9 oz 5.8 g 12 g 0.001 8 lb 0.002 6 lb 0.005 9 lb 0.041 lb	NIST Class F Weights  Orlando, FL
Volumetric Recipients (Graduated Volumetric Containers)	(1 to 20) ml (>20 to 80) ml (>80 to 220) ml	0.06 ml 0.2 ml 0.23 ml	Analytical Balance  Fremont, CA
Pipettes	(1 to 10) $\mu$ L (>10 to 100) $\mu$ L (>100 to 1 000) $\mu$ L	0.051 $\mu$ L 0.079 $\mu$ L 0.17 $\mu$ L	Pipette Calibration Balances  Fremont, CA

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Radiation (Infrared) Thermometers	(50 to 100) $^{\circ}$ C (100 to 300) $^{\circ}$ C (300 to 500) $^{\circ}$ C	1.7 $^{\circ}$ C 5.3 $^{\circ}$ C 8.2 $^{\circ}$ C	Blackbody Source (Flat plate) $\epsilon = 0.95, \lambda = (8 \text{ to } 14) \mu\text{m}$ Vista, CA Fremont, CA
Relative Humidity – Measure <sup>1</sup>	Up to 90 %RH (90 to 95) %RH	1.6 %RH 2.2 %RH	Humidity Probe Monitor  Vista, CA Fremont, CA Orlando, FL
Relative Humidity – Measuring equipment	Up to 90 %RH (90 to 95) %RH	1.9 %RH 2.5 %RH	Humidity chamber, Humidity Probe Monitor  Fremont, CA Orlando, FL

**Thermodynamic**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Measure <sup>1</sup>	(-200 to 0) °C (0 to 100) °C (100 to 300) °C (300 to 420) °C (420 to 650) °C	0.03 °C 0.05 °C 0.07 °C 0.11 °C 0.14 °C	Reference Thermometer w/ PRT  Vista, CA Fremont, CA Orlando, FL
Temperature – Measuring equipment	(-70 to 0) °C (0 to 100) °C (100 to 175) °C	1.3 °C 1.5 °C 2.8 °C	Reference Chamber and Thermometer w/ PRT  Fremont, CA Orlando, FL

**Time and Frequency**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Source <sup>1</sup>	1 μHz to 80 MHz	5.1 x 10 <sup>-12</sup> MHz	Frequency Synthesizer w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL
Frequency – Source <sup>1</sup>	10 MHz to 50 GHz	5.1 x 10 <sup>-12</sup> GHz	Synthesized Sweeper w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL
Frequency – Measure <sup>1</sup>	1 μHz to 12.4 GHz 1 Hz to 50 GHz	5.1 x 10 <sup>-12</sup> GHz 5.1 x 10 <sup>-12</sup> GHz	Electronic Counters, Analyzers w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL

**Time and Frequency**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Time Interval	50 ns to 999 s	5 parts in $10^{12}$ s + 0.43 ns	Universal Counter w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL
Period	4.44 ns to 10 s	0.012 % of reading + 0.17 ns	Universal Counter w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL
Rise/Fall Time - Measure	> 2 ns	0.076 ns	Digital Oscilloscope  Vista, CA Fremont, CA Orlando, FL
Pulse Width - Measure	> 5 ns	1.1 ns	Universal Counter w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL
Stop Watches – Totalize Method	1 s to 24 hr	0.000 6 % of reading + 0.044 s	Universal Counter and Synthesizer w/ GPS Reference  Vista, CA Fremont, CA Orlando, FL

**DIMENSIONAL MEASUREMENT**

**1 Dimensional**

<b>Parameter/Equipment</b>	<b>Range</b>	<b>Expanded Uncertainty of Measurement (+/-)</b>	<b>Reference Standard, Method, and/or Equipment</b>
Distance Measure <sup>3</sup>	Up to 500 in	0.053 in + 0.004 % of reading	Fluke 419D Laser Distance Meter Fremont, CA

**1 Dimensional**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Distance/Linearity Measurement	Up to 610 mm	(19 + 0.7 L/600) μm	Linear Height Gage Fremont, CA

**2 Dimensional**

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional (non-Contact)	X axis to 100 mm Y axis to 80 mm	(14 + 7.5 L/1 000) μm (14 + 7.5 L/1 000) μm	Vision System Fremont, CA

**TESTING**

**Environmental**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
Face Velocity Test Airflow Smoke Pattern Test Leak/Backstreaming Test	ISO 14644-4 ANSI/ASHRAE 110 IEST-RP-CC002.4	Flow hood Devices Test	Thermal Anemometer Aerosol Generator Digital Aerosol Photometer  Vista, CA Fremont, CA Orlando, FL
Airborne Particle Count Survey Airflow Measurement Airflow Smoke Pattern Test HEPA/ULPA filter leak Test Biological Safety Cabinet Classification	ISO 14644-1 ISO 14644-4 IEST-RP-CC034.2 IEST-RP-CC006.3 NSF/ANSI49-2004 Annex	Biological Safety Cabinet Test	Particle Counter Balometer Aerosol Generator Digital Aerosol Photometer Thermal Anemometer  Vista, CA Fremont, CA Orlando, FL

**Environmental**

<b>Specific Tests and/or Properties Measured</b>	<b>Specification, Standard, Method, or Test Technique</b>	<b>Items, Materials or Product Tested</b>	<b>Key Equipment or Technology</b>
Airflow Measurement Air Pressure Difference Test Airflow Smoke Pattern Test HEPA/ULPA filter leak Test Airborne Particle Count Survey Lighting Level Sound Level Test Temperature Test Humidity Test Temperature/Humidity Uniformity Test Recovery Test	ISO 14644-1 ISO 14644-3 ISO 14644-4 IEST-RP-CC006.3 IEST-RP-CC034.2	Clean Room Test	Balometer Differential Pressure Meter Digital Aerosol Photometer, Aerosol Generator Particle Counter Light Level meter Sound meter Temperature & humidity meter Thermal anemometer.  Vista, CA Fremont, CA Orlando, FL
Airborne Particle Count Survey Airflow Velocity Laminar Hood HEPA/ULPA filter leak Test Induction Leak/Backstreaming Test Airflow Smoke Pattern Test Lighting Level Sound Level Test	ISO 14644-1 ISO 14644-4 IEST-RP-CC002.4 IEST-RP-CC006.3 IEST-RP-CC034.2	Laminar Air Flow Workstation Test	Particle Counter Thermal Anemometer Aerosol Generator Digital Aerosol Photometer Light Level meter Sound meter  Vista, CA Fremont, CA Orlando, FL
Oil Aerosol & Vapor Content <sup>4</sup>	ISO 8573-1 ISO 8573-2 ISO 8573-5	Compressed Air Purity Test	Oil Content Analyzer Air Sampler  Vista, CA Fremont, CA Orlando, FL
Humidity / Dew Point Measurement	ISO 8573-3	Compressed Air Purity Test	Dew Point Meter & Diffuser  Vista, CA Fremont, CA Orlando, FL
Pressurized Air Particle Content	ISO 8573-4	Compressed Air Purity Test	Particle Counter & Diffuser  Vista, CA Fremont, CA Orlando, FL


**Environmental**

Specific Tests and/or Properties Measured	Specification, Standard, Method, or Test Technique	Items, Materials or Product Tested	Key Equipment or Technology
<p>Viable Microbiological Contaminant <sup>5</sup></p>	<p>ISO 8573-7</p>	<p>Compressed Air Purity Test</p>	<p>Microbiological Sampler</p> <p>Vista, CA Fremont, CA Orlando, FL</p>

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ( $k=2$ ), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2.  $\Gamma = \rho$
3. Unitless linear measure.
4. Portions of ISO 8573-5 requiring analysis using gas chromatography are contracted to another accredited laboratory.
5. This analysis is intended to be used in conjunction with the testing per ISO 8573-4 when there is a need to identify solid particles that are also viable, colony-forming units and is contracted to another accredited laboratory.
6. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-1736.



Jason Stine, Vice President

